


FlowflexTM



SARS-CoV-2 Antigen Rapid Test

A rapid, highly reliable and affordable kit, providing an aid in early diagnosis of individuals who are suspected of COVID-19 by their healthcare provider.

The King of Rapid Test with **25** Years Experience.



Fast



Reliable



Easy to Use



User friendly



CE Marked

ACON[®]

ACON Biotech (Hangzhou) Co., Ltd.



SARS-CoV-2 Antigen Rapid Test Package Insert

REF L031-11815 English

A rapid test for the qualitative detection of SARS-CoV-2 nucleocapsid antigens in nasal swab specimens.

For professional in vitro diagnostic use only.

INTENDED USE

The SARS-CoV-2 Antigen Rapid Test is a lateral flow chromatographic immunoassay for the qualitative detection the nucleocapsid protein antigen from SARS-CoV-2 in nasal swab specimens directly from individuals who are suspected of COVID-19 by their healthcare provider within the first seven days of the onset of symptoms. The SARS-CoV-2 Antigen Rapid Test does not differentiate between SARS-CoV and SARS-CoV-2.

Results are for the identification of SARS-CoV-2 nucleocapsid antigen. This antigen is generally detectable in upper respiratory samples during the acute phase of infection. Positive results indicate the presence of viral antigens, but clinical correlation with patient history and other diagnostic information is necessary to determine infection status. Positive results do not rule out bacterial infection or co-infection with other viruses. The agent detected may not be the definite cause of disease.

Negative results, from patients with symptom beyond seven days, should be treated as presumptive and confirmed with a molecular assay, if necessary, for patient management. Negative results do not rule out SARS-CoV-2 infection and should not be used as the sole basis for treatment or patient management decisions, including infection control decisions. Negative results should be considered in the context of a patient's recent exposures, history and the presence of clinical signs and symptoms consistent with COVID-19.

The SARS-CoV-2 Antigen Rapid Test is intended for use by trained clinical laboratory personnel and individuals trained in point of care settings.

SUMMARY

The novel coronaviruses belong to the β genus.¹ COVID-19 is an acute respiratory infectious disease. People are generally susceptible. Currently, the patients infected by the novel coronavirus are the main source of infection; asymptomatic infected people can also be an infectious source. Based on the current epidemiological investigation, the incubation period is 1 to 14 days, mostly 3 to 7 days. The main manifestations include fever, fatigue and dry cough. Nasal congestion, runny nose, sore throat, myalgia and diarrhea are found in a few cases.

PRINCIPLE

The SARS-CoV-2 Antigen Rapid Test is a qualitative membrane based chromatographic immunoassay for the qualitative detection of the nucleocapsid protein antigen from SARS-CoV-2 in human nasal swab specimens.

When specimens are processed and added to the test cassette, SARS-CoV-2 antigens, if present in the specimen, will react with the anti-SARS-CoV-2 antibody-coated particles, which have been pre-coated on the test strip. The mixture then migrates upward on the membrane by capillary action. The antigen-conjugate complexes migrate across the test strip to the reaction area and are captured by a line of antibody bound on the membrane. Test results are interpreted visually at 15 minutes based on the presence or absence of visually colored lines. To serve as a procedure control, a colored line will always appear in the control line region indicating that proper volume of specimen has been added and membrane wicking has occurred.

REAGENTS

The test cassette contains anti-SARS-CoV-2 antibodies coated particles on the membrane. The positive control swab contains SARS-CoV-2 recombinant antigen pre-coated on the swab.

PRECAUTIONS

- For professional *in vitro* diagnostic use only. Do not use after the expiration date.
- Do not eat, drink, or smoke in the area where the specimens or kits are handled.
- Do not use the test if the pouch is damaged.
- Handle all specimens as if they contain infectious agents. Observe established precautions against biological hazards throughout testing and follow the standard procedures for proper disposal of specimens.
- Wear protective clothing such as laboratory coats, disposable gloves, and eye protection when specimens are being tested.
- The used test should be discarded according to local regulations. The used test should be considered potentially infectious and be discarded according to local regulations.
- Humidity and temperature can adversely affect results.
- This package insert must be read completely before performing the test. Failure to follow directions in insert may yield inaccurate test results.
- The test line for a high viral load sample may become visible within 15 minutes, or as soon as the sample passes the test line region.
- The test line for a low viral load sample may become visible within 30 minutes.

STORAGE AND STABILITY

- The kit can be stored at temperatures between 2 - 30 °C.
- The test is stable until the expiration date printed on the sealed pouch.
- The test must remain in the sealed pouch until use.
- DO NOT FREEZE.
- Do not use after the expiration date.

MATERIALS

Materials Provided

- Test Cassettes
- Positive Control Swab
- Disposable Swabs*
- Extraction Buffer Tubes
- Negative Control Swab
- Package Insert

* The Disposable Swabs are produced by another manufacturer.


Materials Required But Not Provided

- Personal Protective Equipment
- Timer

SPECIMEN COLLECTION AND PREPARATION

- The SARS-CoV-2 Antigen Rapid Test can be performed using nasal swab specimens.
- Testing should be performed immediately after specimen collection, or at most within one (1) hour after specimen collection, if stored at room temperature (15-30°C).
- To collect a nasal swab sample:

1. Carefully insert a Disposable Swab, **provided with your kit**, into one nostril. Using gentle rotation, push the swab up to 2.5 cm (1 inch) from the edge of the nostril.
2. Rotate the swab 5 times against the mucosa inside the nostril to ensure sufficient specimen collection.




3. Using the same swab, repeat this process in the other nostril to ensure that an adequate amount of sample is collected from both nasal cavities.
4. Withdraw the swab from the nasal cavity. The specimen is now ready for preparation using the extraction buffer tubes.



DIRECTIONS FOR USE

Allow the test and extraction buffer to reach room temperature (15-30 °C) prior to testing.

1. Use an extraction buffer tube for each specimen to be tested and label each tube appropriately.
2. Unscrew the dropper cap from the extraction buffer tube without squeezing.
3. Insert the swab into the tube and swirl it for 30 seconds. Then rotate the swab at least 5 times while squeezing the sides of the tube. Take care to avoid splashing contents out of the tube.
4. Remove the swab while squeezing the sides of the tube to extract the liquid from the swab.
5. Screw the dropper cap firmly onto the extraction buffer tube containing the sample. Mix thoroughly by swirling or flicking the bottom of the tube.
6. Remove the test cassette from the foil pouch and use it as soon as possible.
7. Place the test cassette on a flat and clean surface.
8. Add the processed specimen to the sample well of the test cassette.
 - a. Unscrew the small cap from the dropper tip.
 - b. Invert the extraction buffer tube with the dropper tip pointing downwards and hold it vertically.
 - c. Gently squeeze the tube, dispensing 4 drops of the processed specimen into the sample well.
9. Wait for the colored line(s) to appear. The result should be read at 15-30 minutes. **Do not read the result after 30 minutes.**



INTERPRETATION OF RESULTS

(Please refer to the illustration above)

NEGATIVE: Only one colored control line appears in the control region (C). No apparent colored line appears in the test line region (T). This means that no SARS-CoV-2 antigen was detected.

POSITIVE:* Two distinct colored lines appear. One line in the control line region (C) and the other line in the test line region (T). This means that the presence of SARS-CoV-2 antigen was detected.

***NOTE:** The intensity of the color in the test line (T) may vary depending on the level of the SARS-CoV-2 antigen present in the specimen. Therefore, any shade of color in the test line region (T) should be considered positive.

INVALID: Control line fails to appear. Insufficient specimen volume or incorrect operation are the most likely reasons for control line failure. Review the procedure and repeat the test with a new test cassette. If the problem persists, discontinue using the test kit immediately and contact your local distributor.

QUALITY CONTROL

Internal procedural controls are included in the test. A colored line appearing in the control line region (C) is an internal procedural control. It confirms sufficient specimen volume and correct procedural technique.

Positive and Negative control swabs are supplied with each kit. These control swabs should be used to ensure that the test cassette and that the test procedure is performed correctly. Follow the **"DIRECTIONS FOR USE"** section to perform the control test.

LIMITATIONS

- The SARS-CoV-2 Antigen Rapid Test is for *in vitro* diagnostic use only. The test should be used for the detection of SARS-CoV-2 antigens in nasal swab specimens only. The intensity of the test line does not necessarily correlate to SARS-CoV-2 viral titer in the specimen.
- Specimens should be tested as quickly as possible after specimen collection and at most within the hour following collection.
- Use of viral transport media may result in decreased test sensitivity.
- A false-negative test may result if the level of antigen in a sample is below the detection limit of the test or if the sample was collected incorrectly.
- Test results should be correlated with other clinical data available to the physician.
- A positive test result does not rule out co-infections with other pathogens.
- A positive test result does not differentiate between SARS-CoV and SARS-CoV-2.
- A negative test result is not intended to rule out other viral or bacterial infections.
- A negative result, from a patient with symptom onset beyond seven days, should be treated as presumptive and confirmed with a molecular assay, if necessary, for clinical management.
(If the differentiation of specific SARS viruses and strains is needed, additional testing is required.)

PERFORMANCE CHARACTERISTICS

Clinical Sensitivity, Specificity and Accuracy

The performance of SARS-CoV-2 Antigen Rapid Test was established with 304 nasal swabs collected from individual symptomatic patients (within 7 days of onset) who were suspected of COVID-19. The results show that the relative sensitivity and the relative specificity are as follows:

Clinical Performance for SARS-CoV-2 Antigen Rapid Test

Method	Results	RT-PCR		Total Results
		Negative	Positive	
SARS-CoV-2 Antigen Rapid Test	Negative	269	1	270
	Positive	1	33	34
Total Results		270	34	304

Relative Sensitivity: 97.1% (83.8%-99.9%)*

Relative Specificity: 99.6% (97.7%-99.9%)*

Accuracy: 99.3% (97.5%-99.9%)*

*95% Confidence Intervals

Limit of Detection (LoD)

The LOD of SARS-CoV-2 Antigen Rapid Test was established using limiting dilutions of a viral sample inactivated by gamma irradiation. The viral sample was spiked with negative human nasal sample pool into a serial of concentrations. Each level was tested for 30 replicates. The results show that the LOD is 1.6×10^2 TCID₅₀/mL.

Sample SARS-CoV-2 Concentration	% Positive (Tests)
1.28×10^3 TCID ₅₀ /mL	100% (30/30)
6.4×10^2 TCID ₅₀ /mL	100% (30/30)
3.2×10^2 TCID ₅₀ /mL	100% (30/30)
1.6×10^2 TCID ₅₀ /mL	96.7% (29/30)
8×10^1 TCID ₅₀ /mL	0% (0/30)

Cross-Reactivity and Interference

No cross reactivity was observed with specimens from patients infected with coronavirus-229E, coronavirus-NL63, coronavirus-OC43, coronavirus-HKU1^{1,2}, parainfluenza virus type (Type 1, Type 2, Type 3 and Type 4), Influenza A/B, Human rhinovirus, Human Bocavirus, Human respiratory syncytial virus, Human metapneumovirus, Human adenovirus, Enterovirus, Chlamydia pneumoniae, Haemophilus influenzae, Legionella pneumophila, Mycobacterium tuberculosis, Streptococcus pneumoniae, Streptococcus pyogenes, Bordetella pertussis, Mycoplasma pneumoniae, Candida albicans, MERS-coronavirus, Pneumocystis jirovecii. The SARS-CoV-2 Antigen Rapid Test does not differentiate between SARS-CoV and SARS-CoV-2.

The interfering substances (Whole Blood, Dafenlin Oxymetazoline Hydrochloride Spray, Mometasone Furoate Nasal Spray, Fluticasone Propionate, Physiological Seawater Nasal Cleaner) with a certain concentration have no interference on the test of SARS-CoV-2 Antigen Rapid Test.

PRECISION

Intra-Assay

Within-run precision was determined using 10 replicates of specimens: negative control and SARS-CoV-2 antigen positive controls. The specimens were correctly identified >99% of the time.

Inter-Assay

Between-run precision was determined using 10 independent assays on the same specimen: negative specimen and SARS-CoV-2 antigen positive specimen. Three different lots of the SARS-CoV-2 Antigen Rapid Test were tested using these specimens. The specimens were correctly identified >99% of the time.

BIBLIOGRAPHY

- Shuo Su, Gary Wong, Weifeng Shi, et al. Epidemiology, Genetic recombination, and pathogenesis of coronaviruses. Trends in Microbiology, June 2016, vol. 24, No. 6: 490-502
- Susan R. Weiss, Julian L. Leibowitz, Coronavirus Pathogenesis, Advances in Virus Research, Volume 81: 85-164

Index of Symbols

	Manufacturer		Contains sufficient for <n> tests		Temperature limit
	<i>In vitro</i> diagnostic medical device		Use-by date		Do not reuse
	Consult instructions for use		Batch code		Catalogue number
	Authorized representative in the European Community		Date of manufacture		

Index of Contents


SARS-CoV-2 Antigen	SARS-CoV-2 Antigen
Negative Control Swab	Negative Control Swab
Positive Control Swab	Positive Control Swab
Extraction Buffer Tubes	Extraction Buffer Tubes
Disposable Swabs	Disposable Swabs
SARS-CoV-2 Antigen Rapid Test	SARS-CoV-2 Antigen Rapid Test



ACON Laboratories, Inc.
5850 Oberlin Drive, #340
San Diego, CA 92121, USA
www.aconlabs.com



MDSS GmbH
Schiffgraben 41
30175 Hannover, Germany



English

Specimen Collection Guide - Nasopharyngeal Swabs

How to collect a nasopharyngeal swab sample:

1. Tilt patient's head back 70 degrees. Gently and slowly insert a nasopharyngeal swab, provided with your kit, through the nostril parallel to the palate until resistance is encountered.
2. Gently rub and roll the swab, leaving it in place for several seconds to absorb secretions. If a deviated septum or blockage creates difficulty in obtaining the specimen from one nostril, use the same swab to obtain the specimen from the other nostril.
3. Slowly remove swab while rotating it. The specimen is now ready for preparation using the extraction buffer tubes.

Français

Guide de prélèvement des échantillons - Écouvillons nasopharyngés

Comment recueillir un échantillon d'écouvillon nasopharyngé:

1. Inclinez la tête du patient vers l'arrière de 70 degrés. Insérez doucement et lentement un écouvillon nasopharyngé, fourni avec votre kit, à travers la narine parallèle au palais jusqu'à ce qu'une résistance soit rencontrée.
2. Frottez et roulez doucement l'écouvillon, en le laissant en place pendant plusieurs secondes pour absorber les sécrétions. Si un septum dévié ou un blocage rend difficile l'obtention de l'échantillon à partir d'une narine, utiliser le même écouvillon pour obtenir l'échantillon à partir de l'autre narine.
3. Retirez lentement l'écouvillon tout en le faisant tourner. L'échantillon est maintenant prêt à être préparé à l'aide des tubes tampons d'extraction.

Deutsch

Anleitung zur Probenentnahme - Nasopharyngeale Abstriche

So entnehmen Sie eine nasopharyngeale Abstrichprobe:

1. Neigen Sie den Kopf des Patienten um 70 Grad zurück. Führen Sie einen Nasopharyngealtupfer, den Sie mit Ihrem Kit erhalten haben, sanft und langsam durch das Nasenloch parallel zum Gaumen ein, bis Sie auf Widerstand stoßen.
2. Reiben und rollen Sie den Tupfer sanft und warten Sie ihn einige Sekunden, damit er Sekrete aufnehmen kann. Wenn eine Nasenscheidewandverbiegung oder eine Verstopfung Schwierigkeiten bei der Entnahme der Probe aus einem Nasenloch verursacht, verwenden Sie denselben Tupfer zur Entnahme der Probe aus dem anderen Nasenloch.
3. Entfernen Sie den Tupfer langsam und drehen Sie ihn dabei. Die Probe ist nun bereit für die Vorbereitung mit den Extraktionspufferröhrchen.

Italiano

Guida alla raccolta dei campioni - Tamponi nasofaringei

Come prelevare un campione nasofaringeo con un tampone:

1. Inclinare indietro la testa del paziente di 70 gradi. Inserire con cautela e lentamente un tampone nasofaringeo incluso nel kit nella narice, parallelamente al palato, fino a quando viene incontrata una resistenza.
2. Strofinare e ruotare gentilmente il tampone, lasciandolo in posizione alcuni secondi per assorbire le secrezioni. Se una deviazione o una ostruzione del setto rendono difficile la raccolta del campione da una narice, usare lo stesso tampone per prelevare dall'altra narice.
3. Rimuovere lentamente il tampone mentre lo si fa ruotare. Il campione è ora pronto per la preparazione usando le provette di estrazione.

Español

Guía de recolección de muestras - Bastoncillos nasofaríngeos

Para obtener una muestra de hisopado nasofaríngeo:

1. Incline la cabeza del paciente hacia atrás 70 grados. Delicada y lentamente introduzca un bastoncillo nasofaríngeo (provisto en el kit) a través de la fosa nasal de forma paralela al paladar hasta sentir resistencia.
2. Gire el bastoncillo suavemente y frote, dejándolo en su lugar durante varios segundos para absorber las secreciones. Si hay una desviación u obstrucción del tabique que producen dificultades para obtener la muestra de una fosa nasal, utilice el mismo bastoncillo para obtener la muestra de la otra fosa nasal.
3. Retire lentamente el bastoncillo mientras lo gira. Con esto, la muestra estará lista para prepararse utilizando los tubos de extracción con disolución amortiguadora.

Português

Guia de colheita de amostras - Zaragatoas nasofaríngeas

Como colher uma amostra com uma zaragatoa nasofaríngea:

1. Incline a cabeça do paciente para trás num ângulo de 70 graus. Delicada e lentamente, insira uma zaragata nasofaríngea, fornecida com o seu kit, na narina de forma paralela ao palato até encontrar resistência.
2. Esfregue suavemente e role a zaragatoa, deixando-a no respetivo local durante alguns segundos para absorver as secreções. Se um desvio de septo ou bloqueio criar dificuldade na obtenção da amostra de uma narina, utilize a mesma zaragatoa para obter a amostra da outra narina.
3. Remova lentamente a zaragatoa enquanto a gira. A amostra agora está pronta para a preparação utilizando os tubos de tampão de extração.

Ελληνικά

Οδηγός συλλογής δειγμάτων - Ρινοφαρυγγικοί στυλεοί

Για να συλλέξετε ένα δείγμα ρινοφαρυγγικού επιχρίσματος:


1. Ο ασθενής πρέπει να γείρει πίσω το κεφάλι του σε γωνία 70 μοιρών. Εισάγετε αργά και μαλακά τον ρινοφαρυγγικό στυλεό που παρέχεται με το κιτ σας μέσα στο ρουθούνι παράλληλα με τον ουρανίσκο, έως ότου βρείτε αντίσταση.
2. Τρίψτε ελαφρά και στρίψτε τον στυλεό, κρατώντας τον στο σημείο για αρκετά δευτερόλεπτα ώστε να απορροφήσει τις εκκρίσεις. Σε περίπτωση που υπάρχει στραβό διάφραγμα ή εμπόδιο που δυσκολεύει τη λήψη δείγματος από το ένα ρουθούνι, χρησιμοποιήστε τον ίδιο στυλεό για να λάβετε δείγμα από το άλλο ρουθούνι.
3. Αφαιρέστε τον στυλεό αργά ενώ τον περιστρέφετε. Το δείγμα είναι τώρα έτοιμο για προετοιμασία με τη χρήση των σωληναρίων ρυθμιστικού διαλύματος.

Svenska

Guide för provtagning - Nasofaryngeala provpinnar

Så här samlar du in ett nasofaryngealt prov:

1. Luta patientens huvud bakåt 70 grader. För försiktigt och långsamt in av de medföljande nasofaryngeala provpinnarna i näsborren parallellt med gommen tills motstånd uppstår.
2. Gnugga och rulla försiktigt provpinnen och låt den vara på plats i flera sekunder för att absorbera sekret. Om ett avvikande septum eller en blockering skapar svårigheter att få provet från en näsborre kan du använda samma pinne för att ta provet från den andra näsborren.
3. Ta långsamt bort pinnen medan du vrider den. Provet är nu klart för beredning med hjälp av rören med extraktionsbuffert.



Türkçe

Örnek Toplama Rehberi - Nazofaringeal Çubuklar

Nazofaringeal sürüntü örneği nasıl toplanır:

1. Hastanın başını arkaya doğru 70 derece eğin. Kitiyle birlikte sağlanan nazofaringeal çubuğu, dirençle karşılaşıncaya kadar, damağa paralel olarak burun deliğinden nazik bir şekilde ve yavaşça sokun.
2. Çubuğu nazikçe sürün ve çevirin, salgıları emmesi için birkaç saniye yerinde bırakın. Septum deviasyonu veya tıkanma, bir burun deliğinden numuneyi elde etmede zorluk yaratırsa, numuneyi diğer burun deliğinden almak için aynı çubuğu kullanın.
3. Çubuğu döndürerek yavaşça çıkarın. Numune artık ekstraksiyon tamponu tüpleri kullanılarak hazırlanmaya hazırdır.

Dansk

Prøveindsamlings-vejledning - Nasopharyngeale podningspinde

Sådan indsamles en nasopharyngeal podningsprøve:

1. Vip patientens hoved 70 grader tilbage. Indsæt forsigtigt og langsomt en nasopharyngeal podningspind, der leveres med dit sæt, gennem næseboret parallelt med ganen, indtil der mødes modstand.
2. Gnid og rul forsigtigt podepinden, og lad den sidde i flere sekunder for at absorbere sekreter. Hvis en septumdeviation eller blokering skaber problemer med at få prøven fra det ene næsebor, skal du bruge den samme podepind til at tage prøven fra det andet næsebor.
3. Fjern langsomt podepinden, mens du roterer den. Prøven er nu klar til klar til klargøring ved hjælp af ekstraktionsbuffer-rørene.

Suomeksi

Näytteenkeruuopas - Nenänielupumpulipuikot

Nenänielupumpulipuikkonäytteen ottaminen:

1. Kallista potilaan päätä taakse 70 astetta. Vie paketissa toimitettu nenänielupumpulipuikko varovasti ja hitaasti sieraimen kitalaen suuntaisesti, kunnes kohtaat vastusta.
2. Hankaa ja kierrä pumpulipuikkoa varovasti, jätä se paikoilleen useammaksi sekunniksi, jotta siihen imeytyy eritteitä. Jos nenän väliseinän (septumin) muoto tai tukkeutuminen estää näytteenoton yhdestä sieraimesta, käytä samaa pumpulipuikkoa ottamaan näyte toisesta sieraimesta.
3. Vedä pumpulipuikko hitaasti pois pyörittäen sitä. Näyte on valmis preparoitavaksi käyttäen uuttopuskuripulloja.

Nederlands

Verzamelingsgids voor specimen - Nasofaryngeale uitstrijkjes

Een monster afnemen via de neuskeelholte:

1. Kantelt u het hoofd van de patiënt ongeveer 70 graden naar achteren. Schuift u voorzichtig een wattenstaafje, dat meegeleverd wordt bij uw kit, in een van de neusgaten, totdat u weerstand van het neusgehemelte voelt.
2. Draai en wrijf het wattenstaafje een paar seconden op zijn plaats om afscheiding af te nemen. Als een afwijkend neustussenstuk of verstopping het afnemen van monster uit een neusgat bemoeilijkt, gebruik dan hetzelfde wattenstaafje om het monster in het andere neusgat af te nemen.
3. Haal het wattenstaafje langzaam en met een draaiende beweging uit de neus. Het monster kan nu in het reageerbuisje worden gestopt en worden onderzocht.

Polski

Przewodnik po pobieraniu próbek - Wymazy z nosogardzieli

Sposób pobierania próbek z nosogardzieli:

1. Odchylić głowę pacjenta w tył o 70 stopni. Delikatnie i powoli wsunąć wymaz do nosogardzieli, dostarczony z zestawem, przez nozdrze równoległe do podniebienia, aż do napotkania oporu.
2. Delikatnie pocierać i okręcać wymaz, a następnie pozostawić go na kilka sekund na miejscu, aby wchłonął wydzielinę. Jeśli skrzywiona przegroda lub zator powoduje trudności w uzyskaniu próbki z jednego nozdrza, należy użyć tego samego wymazu do uzyskania próbki z drugiego nozdrza.
3. Powoli wyjąć wymaz, obracając go. Próbką jest teraz gotowa do przygotowania przy użyciu probówek z buforem ekstrakcyjnym.

Hrvatski

Vodič za prikupljanje uzoraka - Vatirani štapići za nazofarinks

Kako prikupiti uzorak brisa iz nazofarinksa:

1. Nagnite glavu pacijenta unazad, za 70 stupnjeva. Nježno i lagano uvlačite priloženi vatirani štapić za nazofarinks u nosnicu, paralelno s nepcem, sve dok ne osjetite otpor.
2. Lagano protrljajte i zakrenite vatirani štapić i ostavite ga na mjestu nekoliko sekundi kako bi upio sekret. Ako iskrivljeni septum ili zapušenost stvara poteškoće za dobivanje uzorka iz jedne nosnice, uporabite isti vatirani štapić kako biste uzeli uzorak iz druge nosnice.
3. Lagano izvucite vatirani štapić istodobno ga rotirajući. Uzorak je sad spreman za pripremu pomoću epruveta s puferom za ekstrakciju.

العربية

دليل جمع العينات - المسحات البلعومية الأنفية

كيفية جمع عينة مسحة البلعوم الأنفي:

1. قم بإمالة رأس المريض للخلف 70 درجة. ادخل المسحة البلعومية الأنفية برفق وبيبطء، المقدمة إليك في مجموعة أدواتك، من خلال فتحة الأنف بشكل موازي للحنك حتى تجد مقاومة.
2. حرك المسحة ولفها برفق، ثم اتركها في مكانها لعدة ثوانٍ لامتصاص الإفرازات. إذا تسبب الحاجز المنحرف أو الانسداد في صعوبة الحصول على العينة من إحدى فتحات الأنف، فاستخدم نفس المسحة للحصول على العينة من فتحة الأنف الأخرى.
3. قم بإزالة المسحة ببطء بينما تقوم بلفها. الآن أصبحت العينة جاهزة للتحضير باستخدام أنابيب الاستخراج العازلة.

Slovenščina

Smernice za odvzem vzorcev - Nazofaringealni brisi

Para obtener una muestra de hisopado nasofaríngeo:

1. Nagnite glavo bolnika za 70 stopinj. Nežno in počasi vstavite nazofaringealni bris, ki je priložen kompletu, skozi nosnico vzporedno z nebom, dokler ne naletite na odpor.
2. Bris nežno podrgnite in zavrtite, tako da ostane nekaj sekund na mestu, da absorbira sekrecijo. Če odstopajoči septum ali blokada povzroča težave pri pridobivanju vzorca iz ene nosnice, uporabite isti bris, da vzamete vzorec iz druge nosnice.
3. Počasi odstranite bris, medtem ko ga vrtite. Vzorec je zdaj pripravljen za pripravo z uporabo ekstrakcijskih puferjskih cevi.



Certificate

No. Q5 042074 0031 Rev. 01

Holder of Certificate: Acon Biotech (Hangzhou) Co., Ltd.

No.210 Zhenzhong Road
 West Lake District
 310030 Hangzhou
 PEOPLE'S REPUBLIC OF CHINA

Facility(ies):

Acon Biotech (Hangzhou) Co., Ltd.
 No.210 Zhenzhong Road, West Lake District, 310030 Hangzhou,
 PEOPLE'S REPUBLIC OF CHINA

Certification Mark:



Scope of Certificate:

**Design and Development,
 Production and Distribution of
 In Vitro Diagnostic Test Kits
 and Related Instruments,
 Lancet and Lancing Device**

Applied Standard(s):

EN ISO 13485:2016
 Medical devices - Quality management systems -
 Requirements for regulatory purposes
 (ISO 13485:2016)
 DIN EN ISO 13485:2016

The Certification Body of TÜV SÜD Product Service GmbH certifies that the company mentioned above has established and is maintaining a quality management system, which meets the requirements of the listed standard(s). See also notes overleaf.

Report No.: SH1910622

Valid from: 2019-07-15

Valid until: 2022-07-14

Date, 2019-07-09

Stefan Preiß
 Head of Certification/Notified Body

Declaration of Conformity

ACON Biotech (Hangzhou) Co., Ltd.
No.210 Zhenzhong Road, West Lake District,
Hangzhou, P.R. China, 310030

We declare under our sole responsibility that the in vitro diagnostic device:

Flowflex SARS-CoV-2 Antigen Rapid Test
REF No.:L031-11815
EAN code: 6921756492175 for Nasal swab
6921756492274 for Nasopharyngeal swab


**classified as Others according to the Annex II of the directive 98/79/EC,
meets all the provisions of the directive 98/79/EC on *in vitro*
diagnostic medical devices which apply to it**

**This declaration is according to Annex III
(excluding Section 6) of the Directive.**

Authorized Representative:
MedNet GmbH
Borkstrasse 10
48163 Muenster, Germany

This Declaration of Conformity is valid until 25 May, 2022.

Signed this 28 day of 9, 2020
in Hangzhou, China



.....
Junny You
International Regulatory Affairs Senior Director
ACON Biotech (Hangzhou) Co., Ltd.



ACON BIOTECH (HANGZHOU) CO., LTD.
No.210 Zhenzhong Road, West Lake District, Hangzhou, P.R. China, 310030
Tel: +86-571-87963569 Fax: +86-571-87963570 E-mail: css@aconlab.com.cn



EUROPEAN COMMISSION
DIRECTORATE-GENERAL FOR HEALTH AND FOOD SAFETY

Public health, country knowledge, crisis management
Health Security and Vaccination

EU health preparedness:

A common list of COVID-19 rapid antigen tests, including those of which their test results are mutually recognised, and a common standardised set of data to be included in COVID-19 test result certificates

Agreed by the Health Security Committee

This document was agreed by the HSC on 17 February 2021

A first update to Annex II was agreed by the HSC on 19 March 2021

A first update to Annex I was agreed by the HSC on 10 May 2021

I. Introduction

Robust testing strategies are an essential aspect of preparedness and response to the COVID-19 pandemic, allowing for early detection of potentially infectious individuals and providing visibility on infection rates and transmission within communities. Moreover, they are a prerequisite to adequate contact tracing to limit the spread through prompt isolation. Also in the context of the circulation of SARS-CoV-2 variants of concern, surge testing in addition to existing testing deployment has proven to be key for controlling and suppressing further spread of the virus.

While the reverse transcription real-time polymerase chain reaction (RT-PCR) assay, which is a nucleic acid amplification test (NAAT) remains the ‘gold standard’ for COVID-19 diagnosis, new tests are rapidly entering the market, allowing faster and cheaper ways to detect ongoing infection. Rapid antigen tests, which detect the presence of viral proteins (antigens), are increasingly being used by Member States as a way of further strengthening countries’ overall testing capacity, particularly in case of limited NAAT capacities or where prolonged testing turnaround times results in no clinical utility.

The Health Security Committee agreed on 17 September 2020 on Recommendations for a common EU testing approach for COVID-19¹, setting out various actions for consideration by countries when updating or adapting their testing strategies. The Recommendations included Member States’ first experiences with rapid antigen tests and their deliberations concerning the settings and situations in which these tests should be used. Since then, the Committee has been discussing the use and application of rapid antigen tests in great depth, and has brought together a wealth of (technical) information on the types of tests used in European countries and the conditions applied.

On 21 January 2021, Member States unanimously agreed on a Council Recommendation setting a common framework for the use of rapid antigen tests and the mutual recognition of COVID-19 test results across the EU². The Council Recommendation called on Member States to agree on three concrete deliverables:

1. **A common list of COVID-19 rapid antigen tests** that are considered appropriate for use in the context of the situations described in the Council Recommendation, that are in line with countries’ testing strategies and that:
 - a. carry CE marking;
 - b. meet the minimum performance requirements of $\geq 90\%$ sensitivity and $\geq 97\%$ specificity; and
 - c. have been validated by at least one Member State as being appropriate for their use in the context of COVID-19, providing details on the methodology and results of such studies, such as the sample type used for validation, the setting

¹ https://ec.europa.eu/health/sites/health/files/preparedness_response/docs/common_testingapproach_covid-19_en.pdf

² <https://data.consilium.europa.eu/doc/document/ST-5451-2021-INIT/en/pdf>

in which the use of the test was assessed, and whether any difficulties occurred as regards the required sensitivity criteria or other performance elements.

2. A selection of rapid antigen tests of which Member States will **mutually recognise the test results for public health measures**.
3. **A common standardised set of data to be included in COVID-19 test result certificates**, further facilitating the mutual recognition of COVID-19 test results.

Based on the information collected by the Health Security Committee, and taking into consideration the current epidemiological situation and the testing strategies and approaches that have been put in place across the EU, this document sets out the three deliverables as agreed by Member States. Its content is prepared based on the criteria set out in the Council Recommendation and considers the relevant recommendations published by the Commission³ and technical guidance issued the European Centre for Disease Prevention and Control (ECDC)⁴ and the World Health Organization (WHO)⁵.

II. Common list of rapid antigen tests

Point 11 of the Council Recommendation of 21 January 2021, calls on Member States to, without prejudice to Directive 98/79/EC, agree on and maintain a common and updated list of COVID-19 rapid antigen tests that are considered appropriate for use in the context of the situations described under point 6 and are in line with countries' testing strategies. Moreover, the antigen tests included in the list should:

- (a) Carry CE marking;
- (b) Meet the minimum performance requirements of $\geq 90\%$ sensitivity and $\geq 97\%$ specificity; and
- (c) Have been validated by at least one Member State as being appropriate for their use in the context of COVID-19, providing details on the methodology and results of such studies, such as the sample type used for validation, the setting in which the use of the test was assessed, and whether any difficulties occurred as regards the required sensitivity criteria or other performance elements.

This list should be shared with ECDC and the Commission to prevent duplication of work and to feed into ongoing initiatives, particularly the "COVID-19 In Vitro Diagnostic Devices and Test Methods Database"⁶, hosted by the Joint Research Centre (JRC). Annex I to this document sets out a common list of rapid antigen tests that meet the criteria as specified above. This list has been incorporated by the JRC in its COVID-19 In Vitro Diagnostic Devices and Test Methods Database. An update to Annex I was agreed by the Health Security Committee on 10 May 2021.

³ <https://eur-lex.europa.eu/legal-content/EN/TXT/?uri=CELEX:32020H1595> and <https://eur-lex.europa.eu/legal-content/EN/TXT/PDF/?uri=CELEX:32020H1743&from=EN>

⁴ <https://www.ecdc.europa.eu/en/publications-data/options-use-rapid-antigen-tests-covid-19-eueea-and-uk>

⁵ <https://www.who.int/publications/i/item/9789240017740>

⁶ <https://covid-19-diagnostics.jrc.ec.europa.eu/devices>

The common list of rapid antigen tests is regularly being reviewed by Member States in the context of Health Security Committee meetings, and, if necessary, be updated in line with new results from independent validation studies becoming available and new tests entering the markets. These updates are also taking into account how mutations of the SARS-CoV-2 virus may affect the efficacy of any particular rapid antigen tests, allowing for the removal of tests no longer deemed effective. The effect of mutations of the SARS-CoV-2 virus on the efficacy of NAAT, in particular RT-PCR assays, will also be kept under review.

III. Rapid antigen tests of which the test results are mutually recognised

As stipulated in point 15 of the Council Recommendation of 21 January 2021, Member States will agree on a selection of rapid antigen tests of which they will **mutually recognise the test results for public health measures**, based on the information included in the common list (see Annex I).

The Health Security Committee agrees that, for rapid antigen test results to be mutually recognised, at least three Member States should be using a rapid antigen tests in practice. Based on this criterion, those rapid antigen tests for which Member States agree that their results will be mutually recognised for public health measures, are highlighted in yellow in Annex I⁷. An update to Annex I, including the selection of tests of which their results are mutually recognised, was agreed by the Health Security Committee on 10 May 2021.

Whenever Member States will review the common list of rapid antigen tests and consider whether any tests should be added or deleted, they will also take into account – also based on new results from independent national validation studies - whether any rapid antigen tests should be removed from or added to the selection of rapid antigen tests of which their results are being mutually recognised. This information will be provided to the JRC, who will update its database accordingly.

IV. Common standardised set of data for COVID-19 test certificates

In order to facilitate in practice the mutual recognition of results of rapid antigen tests as well as NAAT, including RT-PCR assays, point 18 of Council Recommendation 2020/1475 defines that Member States should agree on a common standardised set of data to be included in the form for test result certificates.

Based on information that was submitted by members of the Health Security Committee in response to a survey on mutual recognition on COVID-19 test results and further discussions that took place in the context of the Health Security Committee, Member States agree on the common standardised set of data for COVID-19 test result certificates as presented in Annex II. An update to this Annex was agreed by the Health Security Committee on 19 March 2021, addressing input received from the eHealth Network and in particular the Semantic Subgroup

⁷ This list has been incorporated by the JRC in its COVID-19 In Vitro Diagnostic Devices and Test Methods Database.

and based on discussions that took place in the context of the EU Digital Green Certificate. Member States agree that COVID-19 test results should be made available in the national language(s) of the country where the test was taken, as well as English.

The Health Security Committee will discuss, whenever relevant, possible updates to the agreed common standardised set of data for COVID-19 test certificates, and publish, if necessary, an updated agreed document.

V. Continuous discussions and further work on the common rapid antigen tests list and common dataset for COVID-19 test result certificates

As described in the sections above, the content of this document, as agreed by the Health Security Committee on 17 February 2021, will continue to be discussed by Member States and updated whenever deemed relevant. Whenever updates are required, these will be published as an update to this current document and/or as an update to the JRC COVID-19 In Vitro Diagnostic Devices and Test Methods Database, depending on scope of the update.

Based on the increasing political and commercial interest in the HSC agreed common rapid antigen test list, including those of which their results are mutually recognised by EU Member States, on 21 April 2021, the Commission and JRC presented to the HSC a new procedure for updating the lists. This includes setting up a HSC Technical Working Group on rapid antigen tests, who will play a key role in reviewing the information submitted by EU countries (as well as manufactures) on the use and performance of rapid antigen tests that are available on the market. Once established, the HSC Technical Working Group will, in particular, address the following points:

Common RAT list

➤ Sampling methods to be used

The current HSC agreed common list of rapid antigen tests includes tests for which their clinical performance was measured based on samples collected from nasal, oropharyngeal or nasopharyngeal specimens. Other rapid antigen tests exist that have been validated in EU Member States based on alternative samples, such as saliva, sputum and/or faeces. Further discussions are required to reach consensus on whether these tests should also be included in the HSC agreed common RAT list.

➤ Harmonised methodology for national validation studies on the clinical performance of rapid antigen tests

This will be addressed by future guidelines to be developed by the JRC and the ECDC, also taking into consideration the implementation guide published by WHO on 21 December 2020 on SARS-CoV-2 antigen-detecting rapid diagnostic tests⁸ as well as the guidance that is being developed by the MDCG-IVD Working Group.

⁸ <https://www.who.int/publications/i/item/9789240017740>

Moreover, Member States will continue sharing details via the HSC on the implementation of national validation studies, particularly concerning the validation methodologies and protocols applied.

➤ **Quality of data produced through independent validation studies**

It is key that the sensitivity levels of the rapid antigen tests, as reported by independent national validation studies, reflect clinical performance as measures in practice, rather than the sensitivity reported by the manufacturer. In this context, the JRC is planning to verify the science behind the validation data that has been made available from the Member States through the Health Security Committee, and to verify the findings (eventually in laboratory settings). For the validation of rapid antigen tests, the JRC plans to use the “gold standard” method of NAAT, in particular RT-PCR, by benchmarking the antigen test samples against qPCR and digital PCR.

Moreover, Member States will continue sharing details via the HSC on the results produced by national validation studies, particularly concerning the sample type used for validation, the setting in which the use of the test was assessed, and whether any difficulties occurred as regards the required sensitivity criteria or other performance elements.

➤ **Occurrence of SARS-CoV-2 variants of concern**

Future updates to the common rapid antigen tests list should also take into account how mutations of the SARS-CoV-2 virus may affect the efficacy of any particular rapid antigen tests, allowing for the removal of tests no longer deemed effective. The effect of mutations of the SARS-CoV-2 virus on the efficacy of RT-PCR tests should also be kept under review. In particular, in the current context of circulation of variants of concern, the use of rapid antigen tests does not allow samples to be used for subsequent detection of new variants (by NAAT and/or sequencing).

Mutual recognition of COVID-19 test results

➤ **Criteria to be used for the mutual recognition of rapid antigen test results**

At the moment, the extent to which rapid antigen tests are being used in practice by Member States differs greatly. In this context, Member States have agreed that, for now, the criterion that at least 3 Member States should be using a specific type of rapid antigen test in practice for it to be mutually recognised, applies. Member States will further discuss and explore whether other criteria should be used in the future. It is key that such discussions are held in the context of quality assurance measures.

➤ **Context in which mutual recognition should be applied**

Member States should further discuss the situation in which there is a need for mutual recognition of rapid antigen test results (as well as other COVID-19 test results). In addition to the context of travel, it is relevant to further discuss between countries when the list of rapid antigen tests of which their results will be mutually recognised should be applied.

ANNEX I: Common list of rapid antigen tests⁹

As agreed by Member States on 17 February 2021 and updated on 10 May 2021

The entries highlighted in yellow are the RATs of which Member States have agreed to mutually recognise their test results for public health measures

Manufacturer	RAT commercial name	CE marking	Clinical performance Data by manufacturer	Clinical performance Data used in MS	FIND evaluation studies	EU Member States using in practice	Other countries using in practice	Countries that have completed practical validation studies	MS currently validating	In JRC database (Device ID #) ¹⁰	In FIND database
AAZ-LMB	COVID-VIRO® Rapid antigen test COVID-19	Yes	96.1% sensitivity 100% specificity	BE: 96.6% sensitivity, 100% specificity, NP swab FR: >95% sensitivity, 100% specificity SI: 96.6% sensitivity, 100% specificity, NP swab		BE, FR, SI	CH	FR CH		Yes (1833)	Yes
Abbott Rapid Diagnostics	Panbio™ COVID-19 Ag Rapid Test	Yes	91.4% sensitivity 99.8% specificity Nasal/NP swab	BE: 93.3% sensitivity, 99.4% specificity, NP Swab 98.1% sensitivity, 99.8% specificity, Nasal swab DE: 91.4% sensitivity 99.8% specificity, NP swab 98.1% sensitivity, 99,8 specificity, Nasal swab	FIND Evaluation - Studies in DE and CH, NP swab, 10 Dec 2020	AT, BE, BG, CY, CZ, DE ^[2] , DK, EE, EL, ES, FR ^[1] , HR, IT, LT, LV, MT, NL ^[5] , PL, PT, RO, SE, SK	CH, ME, MK, NO, UK, UA	DE ^[2] , ES, NL ^[5] CH, NO	CY, ES, HR, HU, IE, LU, PT, SE	Yes (1232)	Yes
Acon Biotech (Hangzhou) Co., Ltd	Flowflex SARS-CoV-2 Antigen Rapid Test	Yes	97.1% sensitivity 99.6% specificity Nasal swab	BE: 96.9% sensitivity, 99.5% specificity, NP swab DE: 97.1% sensitivity, 99.5% specificity, NP/Nasal swab	Ongoing	AT, BE, LT, LV, SI		DE^[2]		Yes (1468)	Yes
AESKU.DIAGNOSTICS GmbH & Co, KG	AESKU.RAPID SARS-CoV-2	Yes		DE: 96% sensitivity, 98% specificity SI: 96% sensitivity, 98% specificity, Nasal swab		AT, DE ^[2] , SI		DE ^[2]		No	No

⁹ This is the list of RATs as referred to by the Proposal for a Regulation of the European Parliament and of the Council on a framework for the issuance, verification and acceptance of interoperable certificates on vaccination, testing and recovery to facilitate free movement during the COVID-19 pandemic (Digital Green Certificate), COM/2021/130 final, of 17 March 2021, which is currently being negotiated in the European Parliament and the Council. Member States shall issue and accept Digital Green Certificates based on this list (and subsequent updates).

¹⁰ In case rapid antigen tests are not included in the JRC Database, manufacturers are invited to submit this information here: https://covid-19-diagnostics.jrc.ec.europa.eu/contact/feedback_ant.

Manufacturer	RAT commercial name	CE marking	Clinical performance <i>Data by manufacturer</i>	Clinical performance <i>Data used in MS</i>	FIND evaluation studies	EU Member States using in practice	Other countries using in practice	Countries that have completed practical validation studies	MS currently validating	In JRC database (Device ID #) ¹⁰	In FIND database
Affimedix	TestNOW® - COVID-19 Antigen	Yes		DE: 93.7% sensitivity, 99.2% specificity		DE ^[2]		DE ^[2]		No	No
AMEDA Labordiagnostik GmbH	AMP Rapid Test SARS-CoV-2 Ag	Yes	97.3% sensitivity 100% specificity NP swab 97.3% sensitivity 98.8% specificity Nasal swab	BE: 97.3% sensitivity, 100% specificity, NP swab DE: 97.3% sensitivity, 100% specificity SI: 97.3% sensitivity, 100% specificity, NP swab		AT, BE, BG, DE ^[2] HR, PT, SI	CH, UA	DE ^[2] CH	HR	Yes (1304)	Yes
AmonMed Xiamen Biotechnology Co., Ltd.	COVID-19 Antigen Rapid Test Kit (Colloidal Gold)	Yes	95.05% sensitivity Nasal swab	DE: 98.02% sensitivity , 99.6% specificity		DE ^[2]		DE ^[2]		Yes (1763)	Yes
Anbio (Xiamen) Biotechnology Co., Ltd.	Rapid COVID-19 Antigen-Test (colloidal Gold)	Yes	99.2% sensitivity 100% specificity	DE: 99.27% sensitivity, 100% specificity		AT, DE ^[2]		DE ^[2]		Yes (1822)	No
Anhui DeepBlue Medical Technology Co. Ltd	COVID-19 (SARS-CoV-2) Antigen Test Kit (Colloidal Gold)	Yes		BE: 95% sensitivity, 99% specificity, NP/OP swab DE: 97.1% sensitivity, 99.8% specificity		BE, DE ^[2]	UK	DE ^[2]		Yes (1589 or 1736)	Yes
ArcDia International Ltd	mariPOC SARS-CoV-2	Yes	92.3% sensitivity 100% specificity	FI: Meets the minimum performance requirements – see the report for details.		FI		FI		No	Yes
Asan Pharmaceutical CO., LTD	Asan Easy Test COVID-19 Ag	Yes	94.7% sensitivity 97.7% specificity	DE: 94.67% sensitivity, 97.71% specificity		DE ^[2]		DE ^[2]		Yes (1654)	Yes
Atlas-Link (Beijing) Technology Co. Ltd	NOVA Test ® SARS-CoV-2 Antigen Rapid Test Kit (Colloidal Gold Immunochromatography)	Yes		DE: 97.6% sensitivity, 99.2% specificity		AT, DE ^[2]	CH	DE ^[2] CH		Yes (2010)	Yes
AXIOM Gesellschaft für Diagnostica und Biochemica mbH	COVID-19 Antigen Rapid Test	Yes		DE: 98.1% sensitivity, 100% specificity		DE ^[2]		DE ^[2]		No	No
Azure Biotech, Inc.	Dia Sure COVID-19 Antigen Rapid Test Device	Yes		DE: 94.3% sensitivity, 99.1% specificity		DE ^[2]		DE ^[2]		No	No
Beijing Hotgen Biotech Co., Ltd.	Novel Coronavirus 2019-nCoV Antigen Test (Colloidal Gold)	Yes	97.1% sensitivity 99.76% specificity	BE: 98.6% sensitivity, 100% specificity, NP Swab 97.3% sensitivity, 99.2% specificity. OP swab DE: 95.37% sensitivity, 99.13% specificity SI: 96.6% sensitivity, 99.8% specificity, NP swab	Validation study to start	AT, BE, DE ^[2] , RO, SI		DE ^[2]		Yes (1870)	No

Manufacturer	RAT commercial name	CE marking	Clinical performance <i>Data by manufacturer</i>	Clinical performance <i>Data used in MS</i>	FIND evaluation studies	EU Member States using in practice	Other countries using in practice	Countries that have completed practical validation studies	MS currently validating	In JRC database (Device ID #) ¹⁰	In FIND database
Beijing Lepu Medical Technology	SARS-CoV-2 Antigen Rapid Test Kit (Colloidal Gold immunochromatography)	Yes	92% sensitivity unknown specificity Nasal swab	BE: 92% sensitivity, 99.3% specificity, Nasal DE: 92.0% sensitivity, 99.26% specificity SI: 92% sensitivity, 99.2% specificity, NP swab		AT, BE, DE ^[2] , SI, RO	UA	DE ^[2]		Yes (1331)	Yes
Beijing Wantai Biological Pharmacy Enterprise Co Ltd	WANTAI SARS-CoV-2 Ag Rapid Test (FIA)	Yes	96.6% sensitivity, unknown specificity Nasal swab	DE: 96.6% sensitivity, 96.9% specificity		DE ^[2]		DE ^[2]		Yes (1484)	Yes
BIOSYNEX SWISS SA	BIOSYNEX COVID-19 Ag BSS	Yes		BE: 96% sensitivity, 100% specificity, NP swab DE: 96% sensitivity, 100% specificity		AT, BE, DE ^[2] , DK,FR, NL ^[5] , PT	CH	DE, NL ^[5] , CH		Yes (1223)	Yes
BTNX Inc.	Rapid Response COVID-19 Antigen Rapid Test Device	Yes	94.55% sensitivity 100% specificity	DE: 94.55% sensitivity, 100% specificity		AT, DE ^[2]		DE ^[2]		Yes (1236)	No
CerTest Biotect S.L.	CerTest SARS-CoV-2 CARD TEST	Yes	92.9% sensitivity 99.6% specificity NP swab	BE: 92.9% sensitivity, 99.6% specificity, NP swab SI: 92.9% sensitivity, 98.4% specificity, NP/OP swab		ES, PT, SI		ES		Yes (1173)	Yes
Core Technology Co., ltd	Canea Covid-19 Antigen Rapid Test	Yes		DE: 97.5% sensitivity, 100% specificity		DE ^[2]		DE ^[2]		No	No
Core Technology Co., ltd	Coretests COVID-19 Ag Test	Yes	98.1% sensitivity	DE: 98.1% sensitivity, 99.6% specificity		AT, DE ^[2] , RO		DE ^[2]		Yes (1786)	No
Dialab	DIAQUICK COVID -19 Ag Cassette	Yes		BE: Z20401CE: 93.2% sensitivity, 100% specificity, NP swab Z20601CE: 96.4% sensitivity, 99.2% specificity, NP swab DE: 97.3% sensitivity, 100% specificity		AT, BE, DE ^[2]		DE ^[2]		Yes (1375)	Yes
DDS DIAGNOSTIC	Test Rapid Covid-19 Antigen (tampon nazofaringian)	Yes	98.77% sensitivity 99.03% specificity	RO: Meets the minimum performance requirements.		RO		RO China	RO	Yes (1225)	No

Manufacturer	RAT commercial name	CE marking	Clinical performance <i>Data by manufacturer</i>	Clinical performance <i>Data used in MS</i>	FIND evaluation studies	EU Member States using in practice	Other countries using in practice	Countries that have completed practical validation studies	MS currently validating	In JRC database (Device ID #) ¹⁰	In FIND database
GenBody Inc	GenBody COVID-19 Ag Test	Yes	90% sensitivity 98% specificity NP/OP swab	DE: 90% sensitivity 98% specificity	Withdrawn	DE ^[2]	UA	DE ^[2]		Yes (1244)	Yes
GenSure Biotech Inc	Gensure COVID-19 Antigen Rapid Test Kit (REF: P2004) (DIA-COVID - 19 Ag Rapid Test)	Yes		DE: 96.86% sensitivity, 100% specificity		DE ^[2]		DE ^[2]		Yes (1253)	Yes
Green Cross Medical Science Corp.	GENEDIA W COVID-19 Ag	Yes		BE: 90.2% sensitivity, 100% specificity, NP swab DE: 90.1% sensitivity, 100% specificity		AT, BE, DE ^[2]		DE ^[2]		Yes (1144)	Yes
Guangdong Hecin Scientific, Inc.	2019-nCoV Antigen Test Kit (colloidal gold method)	Yes	96.23% sensitivity 98.51% specificity Nasal swab	DE: 96.6% sensitivity, 99.07% specificity		AT, DE ^[2]		DE ^[2]		Yes (1747)	No
Guangdong Wesail Biotech Co. Ltd	COVID-19 AG Test Kit	Yes	90% sensitivity 98% specificity NP/Nasal swab	DE: 90% sensitivity, 99.2% specificity SI: 90% sensitivity, 98% specificity, NP/Nasal swab		DE ^[2] , SI		DE ^[2]		Yes (1360)	No
Guangzhou Wondfo Biotech Co., Ltd	Wondfo 2019-nCoV Antigen Test (Lateral Flow Method)	Yes		BE: 96.2% sensitivity, 99.7% specificity, NP/OP swab DE: 96.18 % sensitivity, 99.72% specificity		AT, BE, BG, DE ^[2] , FR	CH	DE ^[2]		Yes (1437)	Yes
Hangzhou AllTest Biotech Co., Ltd	ALL TEST Covid 19 Antigen- Rapidtest (Swab)	Yes		AT: 96,4% sensitivity, 99,0% specificity, specimen type: NP; if N sens reduced to: 92,9%		AT		AT	AT	Yes (1256)	Yes
Hangzhou Clongene Biotech Co., Ltd.	COVID-19 Antigen Rapid Test Kit	Yes	98.5% sensitivity unknown specificity Nasal swab	BE: 91.4% sensitivity, 100% specificity, NP/OP swab DE: 91.4% sensitivity, 99.4% specificity SI: 91.4% sensitivity, 100% specificity, NP/OP swab		AT, BE, DE ^[2] , FR, SI	CH	DE ^[2] CH	HR	Yes (1363)	No

Manufacturer	RAT commercial name	CE marking	Clinical performance <i>Data by manufacturer</i>	Clinical performance <i>Data used in MS</i>	FIND evaluation studies	EU Member States using in practice	Other countries using in practice	Countries that have completed practical validation studies	MS currently validating	In JRC database (Device ID #) ¹⁰	In FIND database
Hangzhou Clongene Biotech Co., Ltd.	COVID-19/Influenza A+B Antigen Combo Rapid Test	Yes	91% sensitivity 100% specificity NP swab	DE: 97.7% sensitivity, 99.8% specificity		DE ^[2]		DE ^[2]		Yes (1365)	Yes
Hangzhou Laihe Biotech Co.	LYHER Novel Coronavirus (COVID-19) Antigen Test Kit (Colloidal Gold)	Yes		DE: 96.29% sensitivity, 100% specificity		AT	CH	DE ^[2]		Yes (1215)	No
Hangzhou Lysun Biotechnology Co., Ltd.	COVID-19 antigen Rapid Test Device (Colloidal Gold)	Yes		DE: 96.29% sensitivity, 100% specificity		DE ^[2]	CH	DE ^[2]		No	No
Hangzhou Testsea Biotechnology Co., Ltd.	Testsealabs Covid-19 Antigen Rapid Test Cassette	Yes	92.1% sensitivity 98.1% specificity Nasal swab	DE: 97.6% sensitivity 98.4% specificity		DE ^[2]		DE ^[2]		Yes (1392)	No
Hangzhou Immuno BiotechCo., Ltd	SARS-CoV2 Antigen Rapid Test	Yes		DE: 95.6% sensitivity, 100% specificity		AT, DE ^[2]		DE ^[2]		No	No
Hangzhou Immuno BiotechCo., Ltd	Immunobio SARS-CoV-2 Antigen ANTERIOR NASAL Rapid Test Kit (minimal invasiv)	Yes	94% sensitivity 100% specificity Nasal swab, NP	DE: 94.39% sensitivity 97.67% specificity		DE ^[2]		DE ^[2]		Yes (1844)	No
Healgen Scientific Limited	Coronavirus Ag Rapid Test Cassette (Swab)	Yes		DE: 97.25% sensitivity, 100% specificity SI: 96.7% sensitivity, 99.2% specificity, NP/Nasal swab		AT, DE ^[2] , NL ^[5] , SE, SI	CH	DE ^[2] , NL ^[5]	SE ^[3]	Yes (1767)	No
Humasis Co. Ltd	HUMASIS COVID-19 Ag test	Yes		BE: 95.5% sensitivity, 100% specificity, NP swab DE: 95.5% sensitivity, 100% specificity SI: 95.5% sensitivity, 100% specificity, NP swab		AT, BE, BG, DE ^[2] , FR, HR, SE, SI		DE ^[2]	HR, SE	Yes (1263)	Yes
Joinstar Biomedical Technology	COVID-19 Antigen Rapid Test (Colloidal Gold)	Yes	96.1% sensitivity 98.1% specificity Nasal swab	DE: 96.1% sensitivity, 98.1% specificity SI: 96.1% sensitivity, 98.1% specificity, NP swab		AT, DE ^[2] , PT, SI		DE ^[2]		Yes (1333)	Yes
JOYSBIO (Tianjin) Biotechnology Co., Ltd.	SARS-CoV-2 Antigen Rapid Test Kit (Colloidal Gold immunochromatography)	Yes	98.13% sensitivity	SI: Meets the minimum performance requirements – see the report for details.	FIND evaluation studies in CH 11 Feb 2021	CZ, SI		SI CH		Yes (1764)	Yes

Manufacturer	RAT commercial name	CE marking	Clinical performance <i>Data by manufacturer</i>	Clinical performance <i>Data used in MS</i>	FIND evaluation studies	EU Member States using in practice	Other countries using in practice	Countries that have completed practical validation studies	MS currently validating	In JRC database (Device ID #) ¹⁰	In FIND database
Labnovation Technologies Inc.	SARS-CoV-2 Antigen Rapid Test Kit	Yes		DE: 96.3% sensitivity, 97.3% specificity SI: 96.3% sensitivity, 97.3% specificity, NP/OP swab		DE ^[2] , SI		DE ^[2]		Yes (1266)	Yes
Lumigenex (Suzhou) Co., Ltd	PocRoc SARS-Cov-2 Antigen Schnellnachweiskit (Gold kolloidal)	Yes		DE: 93.33% sensitivity , 99.16% specificity		DE ^[2]		DE ^[2]		No	No
LumiQuick Diagnostics Inc.	QuickProfile™ COVID-19 ANTIGEN Test	Yes		BE: 94% sensitivity, 99% specificity, NP swab DE: 93.7% sensitivity, 98.8% specificity SI: 93.7% sensitivity, 98.8% specificity, NP swab		BE, DE ^[2] , FR, SI,		DE ^[2]		Yes (1267)	Yes
LumiraDX UK LTD	LumiraDx SARS-CoV-2 Ag Test	Yes	97.6% sensitivity 96.7% specificity Nasal swab	DE: 93.8% sensitivity, 98.8% specificity SI: 97.6% sensitivity, 97.7% specificity, NP/Nasal swab		DE ^[2] , ES, SI	CH	DE ^[2] , ES CH		Yes (1268)	No
MEDsan GmbH	MEDsan® SARS-CoV-2 Antigen Rapid Test	Yes	92.5% sensitivity 99.8% specificity NP/OP swab	BE: 92.5% sensitivity, 99.8% specificity, Nasal/OP swab DE: 92.5% sensitivity, 99.8% specificity		AT, BE, DE ^[2]	CH	DE ^[2] CH		Yes (1180)	No
MöLab	COVID-19 Rapid Antigen Test	Yes		DE: 97.25% sensitivity , 99.99% specificity		DE ^[2]		DE ^[2]		Yes (1190)	No
MP Biomedicals Germany	Rapid SARS-CoV-2 Antigen Test Card	Yes	96.39% sensitivity 99.03% specificity Nasal swab	BE: 96.4% sensitivity, 99% specificity, NP/OP swab DE: 96.39 % sensitivity, 99.03% specificity		AT, BE, DE ^[2]	CH	DE ^[2] CH		Yes (1481)	Yes
nal von minden GmbH	NADAL COVID -19 Ag +Influenza A/B Test	Yes		DE: 97.6% sensitivity, 99.9% specificity		DE ^[2]		DE ^[2]		No	No

Manufacturer	RAT commercial name	CE marking	Clinical performance <i>Data by manufacturer</i>	Clinical performance <i>Data used in MS</i>	FIND evaluation studies	EU Member States using in practice	Other countries using in practice	Countries that have completed practical validation studies	MS currently validating	In JRC database (Device ID #) ¹⁰	In FIND database
nal von minden GmbH	NADAL COVID -19 Ag Test	Yes	97.6% sensitivity 99.9% specificity Nasal swab	BE: 97.6% sensitivity, 99.9% specificity, NP/OP swab DE: 97.6% sensitivity, 99.9% specificity SI: 97.6% sensitivity, 99.9% specificity, NP/OP swab	FIND Evaluation studies 26 April 21	AT, BE, DE ^[2] , PT, SI		DE ^[2] , FR China	HR	Yes (1162)	No
NanoEntek	FREND Covid-19 Ag	Yes	94.12% sensitivity 100% specificity NP swab	DE: 94.12% sensitivity , 100% specificity		DE ^[2]		DE ^[2]		Yes (1420)	Yes
Oncosem Onkolojik Sistemler San. ve Tic. A.S.	CAT	Yes	93.75% sensitivity 98.04% specificity Nasal swab	DE: 96.36% sensitivity, 98.04% specificity		DE ^[2]		DE ^[2]		Yes (1199)	No
PCL Inc	PCL COVID19 Ag Rapid FIA	Yes		DE: 94,92 % sensitivity, 99,99 % specificity SI: 95.5% sensitivity, 98.6% specificity, NP/OP swab, sputum		FR, DE, RO, SI		DE ^[2]		Yes (308)	No
PerGrande Biotech Development Co., Ltd.	SARS-CoV-2 Antigen Detection Kit (Colloidal Gold Immunochromatographic assay)	Yes		DE: 94.28% sensitivity, 99.11% specificity		AT, DE ^[2]		DE ^[2]		No	No
Precision Biosensor Inc (Axon Lab SG)	Exdia COVI-19 Ag Test	Yes	93.9% sensitivity 98% specificity NP swab	DE: 93.88% sensitivity , 98% specificity SI: 93.9% sensitivity, 98% specificity, NP swab		SI, DE ^[2]	CH	DE ^[2] CH		Yes (1271)	Yes
Qingdao Hightop Biotech Co Ltd	SARS-CoV-2 Antigen Rapid Test	Yes	95% sensitivity unknown specificity Nasal swab	DE: 95% sensitivity 99.75% specificity		AT, DE ^[2]		DE ^[2]		Yes (1341)	No
Quidel Corporation	Sofia 2 SARS Antigen FIA	Yes	96.7% sensitivity 100% specificity NP/Nasal swab	BE: 96.7% sensitivity, 100% specificity, NP/nasal swab DE: 96.7% sensitivity , 100% specificity SI: 96.7% sensitivity, 100% specificity, NP/Nasal swab		AT, BE, DE ^[2] , FI, NL ^[5] , PT, SI	CH	DE ^[2] , NL ^[5] CH	SI	Yes (1097)	Yes

Manufacturer	RAT commercial name	CE marking	Clinical performance <i>Data by manufacturer</i>	Clinical performance <i>Data used in MS</i>	FIND evaluation studies	EU Member States using in practice	Other countries using in practice	Countries that have completed practical validation studies	MS currently validating	In JRC database (Device ID #) ¹⁰	In FIND database
Rapigen Inc.	BIOCREDIT COVID-19 Ag - SARS-CoV 2 Antigen test	Yes	90.2% sensitivity 100% specificity NP swab	SI: 90.2% sensitivity, 100% specificity, NP swab		AT, RO, SK, FR, SI		HU	PT	Yes (1606)	Yes
Roche (SD BIOSENSOR)	SARS-CoV-2 Antigen Rapid Test	Yes	96.52% sensitivity 99.2% specificity NP	DE: 96.52% sensitivity, 99.68% specificity		AT, DE ^[2] , NL, RO	CH, NO	DE ^[2]		Yes (1604)	Yes
Safecare Biotech Hangzhou Co	COVID-19 Antigen Rapid Test Kit (Swab)	Yes	97.04% sensitivity unknown specificity Nasal swab	DE: 97.27 % sensitivity , 99.42% specificity		AT, DE ^[2] , FR	CH	DE ^[2]		Yes (1489)	No
Safecare Biotech Hangzhou Co	Multi-Respiratory Virus Antigen Test Kit (Swab) (Influenza A+B/COVID-19)	Yes	Sensitivity: 97.04%	DE: 97.04% sensitivity , 99.44% specificity		DE ^[2]		DE ^[2]		Yes (1490)	No
SD BIOSENSOR, Inc.; Roche	STANDARD F COVID-19 Ag FIA	Yes	94,09% sensitivity 98.52% specificity	BE: 96.5% sensitivity, 99.7% specificity, NP swab DE: 94% sensitivity 97% specificity	FIND Evaluation - Studies in DE and Brazil, 10 Dec 2020	AT, BE, BG, DE ^[2] , IT, LU, LV, NL ^[5] , PT, RO, SK	CH	DE ^[2] , IT, NL ^[5] , DK, CH, UK, BR	LU, PT	Yes (344)	Yes
SD BIOSENSOR, Inc.; Roche	STANDARD Q COVID-19 Ag Test	Yes	96.52% sensitivity 99.68% specificity NP swab	BE: 96.5% sensitivity, 99.7% specificity, NP swab DE: 96.52% sensitivity, 99.68% specificity SI: 96.5% sensitivity, 99.7% specificity, NP swab	FIND Evaluation - Studies in DE, CH and Brazil, 10 Dec 2020	AT, BE, BG, CY, DE ^[2] , DK, EE, ES, FI, FR, HR, IT, LU, LV, MT, NL ^[5] , RO, SE, SK, SI	ME, NO, CH	DE ^[2] , ES, IT, NL ^[5] , DK, CH, UA, UK, BR	HR, IE, LU, SI, SE	Yes (345)	Yes
SGA MÜHENDİSLİK DANIŞMANLIK EĞİTİM İÇ VE DIŞ TİC. A.Ş.	V-Chek SARS-CoV2- Rapid Ag Tets (Coloidal Gold)	Yes	96.6% sensitivity, Nasal swab	DE: 96.6% sensitivity, 99% specificity		DE ^[2]		DE ^[2]		Yes (1319)	No
Shenzhen Ultra-Diagnostics Biotech Co.	SARS-COV-2 Antigen test Kit (colloidal gold)	Yes		BE: 92% sensitivity, 100% specificity, NP swab 100% sensitivity, 100% specificity, OP swab 96% sensitivity, 100% specificity, Saliva SI: 95.9% sensitivity, 99.9% specificity, NP/OP/Nasal swab, saliva		AT, BE, ES, SI		BE, SI		No	No
Shenzhen Lvshiyuan Biotechnology Co., Ltd.	Green Spring SARS-CoV-2-Antigen-Schnelltests-Set	Yes		DE: 98% sensitivity , 100% specificity		DE ^[2]		DE ^[2]		No	Yes

Manufacturer	RAT commercial name	CE marking	Clinical performance Data by manufacturer	Clinical performance Data used in MS	FIND evaluation studies	EU Member States using in practice	Other countries using in practice	Countries that have completed practical validation studies	MS currently validating	In JRC database (Device ID #) ¹⁰	In FIND database
Shenzhen Watmind Medical Co., Ltd	SARS-CoV-2 Ag Diagnostic Test Kit (Colloidal Gold)	Yes	95.15% Sensitivity Nasal swab, Salvia	DE: 95.15% sensitivity , 99.12% specificity		AT, DE ^[2] , FR		DE ^[2]		Yes (1769)	No
Shenzhen Zhenrui Biotech Co., Ltd	Zhenrui [®] COVID-19 Antigen Test Cassette	Yes	96% sensitivity Nasal swab, Salvia	DE: 96% sensitivity 97% specificity		DE ^[2]		DE ^[2]		Yes (1574)	No
Siemens Healthineers	CLINITEST Rapid COVID-19 Antigen Test	Yes	96.72% sensitivity 96.72% specificity Nasal swab	BE: 98.32% sensitivity, 99.6% specificity, NP swab 97.25% sensitivity, 100% specificity, Nasal swab SI: 96.7% sensitivity, 99.2% specificity, NP/Nasal swab		AT, BE, DE ^[2] , FR, HR, NL ^[5] , PT, SE, SI	CH	DE ^[2] , ES, NL ^[5]	HR, PT, SE ^[3]	Yes (1218)	Yes
Sugentech, Inc.	SGTi-flex COVID-19 Ag	Yes		DE: 95.1% sensitivity, 99% specificity		AT, DE ^[2]		DE ^[2]		Yes (1114)	No
TODA Pharma	TODA CORONADIAG Ag [®]	Yes	98.6% sensitivity unknown specificity Nasal swab	BE: 96.6% sensitivity, 100% specificity, NP/OP swab DE: 96.6% sensitivity, 100 specificity SI: 96.6% sensitivity, 100% specificity, NP/OP swab		BE, DE ^[2] , SI		DE ^[2]		Yes (1466)	No
Tody Laboratories Int.	Coronavirus (SARS-CoV 2) Antigen - Oral Fluid	Yes	90.1% sensitivity 99.3% specificity	RO: Meets the minimum performance requirements.		RO		ES UA, China	RO	No	Yes
Vitrosens Biyoteknoloji Ltd. Şti.	RapidFor SARS-CoV-2 Ag Test Kit	Yes	97.3% sensitivity unknown specificity Nasal swab, saliva	DE: 97.3% sensitivity, 99% specificity SI: 97.3% sensitivity, 99% specificity, NP/OP/Nasal swab		DE ^[2] , SI		DE ^[2]		Yes (1443)	Yes
VivaChek Biotech (Hangzhou) Co., Ltd.	VivaDiagTM Pro SARS- CoV-2 Ag Rapid Test	Yes		AT: 97,06% sensitivity, 100% specificity, all specimen types, i.e. N&OP&NP swab		AT		AT	AT	Yes (1246)	Yes
Wuhan EasyDiagnosis Biomedicine Co., Ltd.	Antigen-Testkit für COVID- 19 (SARS-Cov-2)	Yes		DE: 96.15% sensitivity , 99.26% specificity		DE ^[2]		DE ^[2]		No	Yes

Manufacturer	RAT commercial name	CE marking	Clinical performance <i>Data by manufacturer</i>	Clinical performance <i>Data used in MS</i>	FIND evaluation studies	EU Member States using in practice	Other countries using in practice	Countries that have completed practical validation studies	MS currently validating	In JRC database (Device ID #) ¹⁰	In FIND database
Xiamen Bosen Biotech Co	Rapid SARS-CoV-2 Antigen Test card	Yes	Not specified	BE: 93.8% sensitivity, 100% specificity, NP swab DE: 96.49% sensitivity, 99.03% specificity		AT, BE, BG, DE ^[2] , FR, RO	CH	DE ^[2] CH		Yes (1278)	Yes
Xiamen Wiz Biotech Co., Ltd.	SARS-CoV-2 Antigen Rapid Test	Yes		DE: 96.3% sensitivity, 100% specificity		AT, DE ^[2]		DE ^[2]		No	Yes
Xiamen Wiz Biotech Co., Ltd.	SARS-CoV-2 Antigen Rapid Test (Colloidal Gold)	Yes		DE: 95.91% sensitivity, 100% specificity		AT, DE ^[2]		DE ^[2]		No	No
Zhejiang Anji Saianfu Biotech Co., Ltd.	AndLucky COVID-19 Antigen Rapid Test	Yes	95.8% sensitivity, Nasal swab, Saliva	DE: 97.5% sensitivity, 99.1% specificity		AT, DE ^[2]		DE ^[2]		Yes (1296)	No
Zhejiang Anji Saianfu Biotech Co., Ltd.	reOpenTest COVID-19 Antigen Rapid Test	Yes	95.8% sensitivity, Nasal swab, Saliva, Plasma	DE: 95.8% sensitivity, 99% specificity		DE ^[2]		DE ^[2]		Yes (1295)	No
Zhejiang Orient Gene Biotech Co., Ltd	Coronavirus Ag Rapid Test Cassette (Swab)	Yes	96.72% sensitivity unknown specificity Nasal swab	BE: 98.32% sensitivity, 99.6% specificity, NP swab 97.25% sensitivity, 100% specificity, Nasal swab DE: 96.72% sensitivity, 99.22% specificity		AT, BE, BG, DE ^[2] , PT	CH, UK	DE ^[2]	SE ^[3]	Yes (1343)	No

Notes:

[1] FR: Reference to validation study (not specifying which specific RAT is being recommended or was tested in practice): https://www.has-sante.fr/upload/docs/application/pdf/2020-10/synthese_tests_antigeniques_vd.pdf

[2] DE: Rapid antigen tests that have completed practical validation studies in Germany: See: https://www.pei.de/SharedDocs/Downloads/DE/newsroom/dossiers/evaluierung-sensitivitaet-sars-cov-2-antigentests-04-12-2020.pdf?__blob=publicationFile&v=43

[3] SE: Smaller evaluations ongoing in some of the regions.

[4] BE: In the clinical performance study performed in three different clinical laboratories during the ascendant phase of the epidemiological curve, we found an overall sensitivity and specificity of 57.6 and 99.5%, respectively with an accuracy of 82.6%.

[5] NL: Collected validation data from accredited laboratories in the Netherlands. The report includes evaluations of various RAT that labs performed at their own initiative. <https://lci.rivm.nl/antigeensneltesten>

ANNEX II: Common standardised set of data to be included in COVID-19 test result certificates, as agreed by Member States on 17 February 2021 and updated on 19 March 2021

Section	Data element	Description	Preferred Code System
Person identification	Person name	The legal name of the tested person. Surname(s) and forename(s), in that order.	
	Person identifier <i>(optional)</i>	An identifier of the tested person, according to the policies applicable in each country. Examples: citizen ID and/or document number (ID-card/passport).	
	Person date of birth <i>(optional)</i>	Tested person's date of birth. Mandatory if no Person identifier is provided.	Complete date, without time, following the ISO 8601.
Test information	Disease or agent targeted	Specification that it concerns the detection of SARS-CoV-2 infection.	ICD-10, SNOMED CT
	Type of test	Description of the type of test that was conducted, e.g. NAAT or rapid antigen test.	LOINC, NPU
	Test name <i>(optional for NAAT)</i>	Commercial or brand name of the test.	
	Test Manufacturer <i>(optional for NAAT)</i>	Legal manufacturer of the test.	
	Sample origin <i>(optional)</i>	The type of sample that was taken (e.g. nasopharyngeal swab, oropharyngeal swab, nasal swab, saliva).	SNOMED CT
	Date and time of the test sample collection	Date and time when the sample was collected.	Complete date, with time and time zone, following ISO 8601
	Date and time of the test result production <i>(optional)</i>	Date and time when the test result was produced.	Complete date, with time and time zone, following ISO 8601
	Result of the test	For example, negative, positive, inconclusive or void.	SNOMED CT
	Testing centre or facility <i>(mandatory for NAAT)</i>	Name/code of testing centre, facility or a health authority responsible for the testing event. <i>Optional:</i> address of the testing facility.	
	Health Professional identification <i>(optional)</i>	Name or health professional code responsible for conducting (and validating) the test. Surname(s) and forename(s), in that order.	
	Country where the test was taken	The country in which the individual was tested.	ISO 3166 Country Codes
Test certificate metadata	Test result certificate issuer	Entity that issued the COVID-19 test result certificate (allowing to check the certificate).	
	Certificate identifier	Reference of the COVID-19 test result certificate (unique identifier).	



Risk Management Report

CoV g-03-06



SARS-CoV-2 Antigen Rapid Test

Revision	Description	By	Date
A	Initial release	Diana Zhang	2020.09.01



Risk Management Report

Table of Contents

1.	Introduction	3
2.	Overview	3
3.	Intended Use/Purpose	3
4.	Scope	3
5.	Software Safety Classification	3
6.	Definitions	6
7.	Failure Mode and Effects Analysis Table.....	6
8.	Signature Page.....	12



Risk Management Report

1. Introduction

The SARS-CoV-2 Antigen Rapid Test is a qualitative detection based on lateral immunoassay for SARS-CoV-2 antigen.

2. Overview

This document provides a safety risk management for SARS-CoV-2 Antigen Rapid Test.

3. Intended Use/Purpose

1) Medical purpose:

Qualitative detection of SARS-CoV-2 antigen.

For professional in vitro diagnostic use only.

2) Part of the body, type of tissue applied to or interacted with, or sample type:

This product is intended to test nasal swab specimens.

3) Operator profile:

Read the package insert carefully before the test. No special training is required.

4) Application:

Add 4~5 drops (approximately 100~125 ul) samples onto the Specimen well of the test cassette, and then start the timer. Wait for the colored line(s) to appear. Read results at 15~20minutes. Allow the test, the specimen should reach room temperature (15-30°C) prior to testing.

4. Scope

This risk management report addresses the safety risks that may affect the patient or the operator as associated with the operation of the SARS-CoV-2 Antigen Rapid Test.

5. Software Safety Classification

There is no software related with this product.

6. Definitions:

i. Severity Safety Classification:

Severity Ranking	Severity Ranking (Descriptive)	S/W Safety Classification	Description
1	Negligible	Class A	Inconvenience or temporary discomfort; no injury to user / operator.



Risk Management Report

2	Minor	Class B	Results in temporary injury or impairment to user / operator. Not requiring professional medical intervention.
3	Serious	Class C	Results in injury or impairment requiring professional medical intervention.
4	Critical		Results in permanent impairment or life-threatening injury.
5	Catastrophic		Results in patient death.

ii. **Probability of Occurrence**

Severity Ranking	Likelihood	Description	Probability Range
1	Improbable	Almost impossible to occur	$< 10^{-6}$
2	Remote	Unlikely to occur	$< 10^{-5}$ and $\geq 10^{-6}$
3	Occasional	Unlikely but possible to occur sometime	$< 10^{-4}$ and $\geq 10^{-5}$
4	Probable	Likely to occur	$< 10^{-3}$ and $\geq 10^{-4}$
5	Frequent	Likely to occur several times	$> 10^{-3}$

iii. **Risk Assessment Table:**

		Risk Assessment				
Probability	Frequent					
	Probable					
	Occasional					
	Remote					
	Improbable					
		Negligible	Minor	Serious	Critical	Catastrophic
		Severity				



Risk Management Report

iv. Risk Assessment Levels:

Level	Risk Acceptability	Description
Grey	Unacceptable Risk	Risk in this region is not tolerated Resolution Required – Redesign, Do not release
Clear	Acceptable risk	Risk is considered to be negligible compared to the risk of other hazards.

Risk Management Report

7. Failure Mode and Effects Analysis Table:

For Risk analysis it is recommended to consider each of these potential sources of error, as applicable to the device, and also consider any other potential system failures that may be specific to the device.

Potential source of Error				Adverse effect	Risk Assessment before mitigation			Risk Control Measure/Mitigation	Risk Assessment after mitigation			Supporting Documents (Risk control Measure/Risk Mitigation)
Risk Factor	Possible Error	Cause	Hazard		Risk Definition		Risk Evaluation		Risk Definition		Risk Evaluation	
					Severity	Probability of Occurrence			Severity	Probability of Occurrence		
	Use of incorrect specimen type	Use of specimen other than nasal swab	Wrong test result	Misdiagnosis	Serious	Occasional	Unacceptable	Product labeling instruct user to use appropriate specimen type	Serious	Improbable	Acceptable	Package Insert
	Incorrect application of the specimen or buffer to the device	Insufficient specimen /buffer or too much specimen/buffer	Wrong test result	Misdiagnosis	Serious	Occasional	Unacceptable	1.Product labeling instruct user to get enough specimen; 2.A control line is set to ensure enough specimen 3. Sample volume flex study was performed in validation study report	Serious	Improbable	Acceptable	Package Insert Validation Study Report, sample and buffer volume flex study
		Incorrect placement specimen	Wrong test result	Misdiagnosis	Serious	Occasional	Unacceptable	Product labeling instruct user how to apply specimen	Serious	Improbable	Acceptable	Package Insert

QUALITY ASSURANCE

Risk Management Report

Operator error/ Human factors	Incorrect handling of reagents	Device pouch was opened for long time before test	Wrong test result	Misdiagnosis	Serious	Occasional	Unacceptable	Product labeling instruct user to use the test as soon as possible after the sealed pouch is opened	Serious	Improbable	Acceptable	Package Insert Validation Study Report, Open pouch study
		Control swab pouch was opened for long time before test	Wrong test result	Misdiagnosis	Serious	Occasional	Unacceptable	Product labeling instruct user to use the test as soon as possible after the sealed pouch is opened	Serious	Improbable	Acceptable	Package Insert Validation Study Report, Open pouch study
	Incorrect placement of device	The device doesn't lay on a flat surface after specimen applying	Wrong test result	Misdiagnosis	Serious	Probable	Unacceptable	Product labeling instruct user to lay the device on a flat surface before specimen applying	Serious	Improbable	Acceptable	Package Insert
	Incorrect timing of procedures	Read result before the time required	Wrong test result	Misdiagnosis	Serious	Occasional	Unacceptable	1. Time flex study was performed 2. Product labeling instruct user to read the result at appropriate time interval	Serious	Improbable	Acceptable	Package Insert Validation Study Report, time flex study
		Read result after the time required	Wrong test result	Misdiagnosis	Serious	Occasional	Unacceptable	1. Time flex study was performed 2. Product labeling instruct user to read the result at appropriate time interval	Serious	Improbable	Acceptable	Package Insert Validation Study Report, time flex study
	Incorrect reading of test results	Confuse the control line and the test line	Wrong test result	Misdiagnosis	Serious	Probable	Unacceptable	Product labeling instruct user to read the appropriate result.	Serious	Improbable	Acceptable	Package Insert

QUALITY ASSURANCE

Risk Management Report

								Product labeling add the graphic result display.				
	Incorrect reading due to color blindness	Lose sight of the low positive line	Wrong test result	Misdiagnosis	Serious	Probable	Unacceptable	Product labeling instruct user to read the appropriate result. Product labeling add the graphic result display.	Serious	Improbable	Acceptable	Package Insert
	Use of polluted device or control swab	Reuse of device	Wrong test result	Misdiagnosis	Serious	Probable	Unacceptable	Product labeling instruct user not to reuse	Serious	Improbable	Acceptable	Package Insert
		Reuse of control swab	Wrong test result	Misdiagnosis	Serious	Probable	Unacceptable	Product labeling instruct user not to reuse	Serious	Improbable	Acceptable	Package Insert
Specimen Integrity and Handling	Error in specimen handling	The swab is not swirled in the fluid	Wrong test result	Misdiagnosis	Serious	Probable	Unacceptable	Product labeling instruct user to swirl the swab in the fluid	Serious	Improbable	Acceptable	Package Insert
		The sample temperature doesn't reach to room temperature prior to testing	Wrong test result	Misdiagnosis	Serious	Probable	Unacceptable	Product labeling instruct user to bring specimens to room temperature prior to testing	Serious	Improbable	Acceptable	Package Insert
	Use of inappropriate swab	swab interference	Wrong test result	Misdiagnosis	Serious	Probable	Unacceptable	Product labeling instruct user to use the appointed swab	Serious	Improbable	Acceptable	package Insert

QUALITY ASSURANCE

Risk Management Report

	Incorrect specimen transport and/or storage.	The specimen is not stored in appropriate temperature and time	Wrong test result	Misdiagnosis	Serious	Probable	Unacceptable	Product labeling instruct user to store the specimen in appropriate temperature and time	Serious	Improbable	Acceptable	Package Insert Sample stability study
	Presence of interfering substances	The test is interfered by other antigens in specimen	Wrong test result	Misdiagnosis	Serious	Probable	Unacceptable	Product design to eliminate the interference from other disease.	Serious	Improbable	Acceptable	Validation Study Report, cross reaction study
Reagent integrity	Use of Improperly Stored Test reagent	The test is damaged by high temperature	Wrong test result	Misdiagnosis	Serious	Probable	Unacceptable	Product labeling instruct user to store the test in the appropriate temperature	Serious	Improbable	Acceptable	Package Insert Stability study
		The test is damaged by high humidity	Wrong test result	Misdiagnosis	Serious	Probable	Unacceptable	Product designed to packaged with a desiccant	Serious	Improbable	Acceptable	Package Insert
	Use of Outdated/ Expired reagents	The product is Outdated	Wrong test result	Misdiagnosis	Serious	Probable	Unacceptable	Product labeling instruct user to run the test before the expiration date.	Serious	Improbable	Acceptable	Package Insert

QUALITY ASSURANCE

Risk Management Report

<p>Environmental Factors</p>	<p>Impact of key environmental factors (heat, humidity, barometric pressure changes, altitude (if applicable), sunlight, surface angle, device movement, etc.) on reagents, specimens, and test results.</p>	<p>The environmental temperature is too low or too high</p>	<p>Wrong test result</p>	<p>Misdiagnosis</p>	<p>Serious</p>	<p>Probable</p>	<p>Unacceptable</p>	<p>Product labeling instruct user to run the test in appropriate environmental temperature</p>	<p>Serious</p>	<p>Improbable</p>	<p>Acceptable</p>	<p>Validation Study Report, temperature flex study</p>
<p>Product package factors</p>	<p>Impact of pouch on testing results such as material and integrity, or impact of product appearance shape on pouch integrity</p>	<p>The product was damaged when transporting</p>	<p>Wrong test result</p>	<p>Misdiagnosis</p>	<p>Serious</p>	<p>Probable</p>	<p>Unacceptable</p>	<p>Shipping study was performed to validate the packaging</p>	<p>Serious</p>	<p>Improbable</p>	<p>Acceptable</p>	<p>Shipping study report</p>
	<p>Impact of pouch on testing results such as material and integrity, or impact of product appearance shape on pouch integrity</p>	<p>The desiccant is missing or disabled in the pouch</p>	<p>Wrong test result</p>	<p>Misdiagnosis</p>	<p>Serious</p>	<p>Probable</p>	<p>Unacceptable</p>	<p>Double check the desiccant in the manufacture. Raw material inspection.</p>	<p>Serious</p>	<p>Improbable</p>	<p>Acceptable</p>	<p>C-0828 Packaging SOP</p>



Risk Management Report

8. Signature Page

The Risk Management Plan has been effectively implemented for this device. The Risk Control Measures' effectiveness and verification methods have been reviewed and all hazards and risks have been reduced to acceptable levels.

Product Identification: SARS-CoV-2 Antigen Rapid Test

Team Leader: Jianxi Kong

Risk Analysis Team Membership

Date: 2020.09.09

Attendees:	Printed Name	Signature	Department	Affiliation
	Jordan Chen	Jordan Chen	MF	Acon bio
	Sihong Zhang	Sihong Zhang	QC	Acon bio
	Eren Jiang	Eren Jiang	Quality	Acon bio
	Lily Fan	Lily Fan	Int'l RA	Acon bio
	Demard Tao	Demard Tao	IP	Acon bio
	Nancy	Nancy	MTI	Acon bio

Design Review Committee Disposition

Department	Recommend Approval	Printed Name	Signature	Date
R & D	<input checked="" type="checkbox"/> Yes [] No [] Cond	Tao Shang	Tao Shang	2020.9.9
Manufacturing	<input checked="" type="checkbox"/> Yes [] No [] Cond	Winston Chen	Winston Chen	2020.9.9
QC	<input checked="" type="checkbox"/> Yes [] No [] Cond	Andy Xu	Andy Xu	2020.9.9
Regulatory	<input checked="" type="checkbox"/> Yes [] No [] Cond	Hysa Lu	Hysa Lu	2020.9.11
Marketing	<input checked="" type="checkbox"/> Yes [] No [] Cond	William Chuy	William Chuy	2020.9.9

Risk Analysis Report Final Disposition (The final disposition is ruled by majority vote of the Design Review Committee. This portion is filled out by the Regulatory representative.)

- Approved
- Approved, but requires follow-up (see comments)
- Not approved, requires additional work (see comments)

Comments: _____

SAFETY DATA SHEET

SECTION 1: IDENTIFICATION OF THE SUBSTANCE/MIXTURE AND OF THE COMPANY/ UNDERTAKING

1.1 Product identifier

Product name: SARS-CoV-2 Antigen Rapid Test

1.2 Relevant identified uses of the substance or mixture and uses advised against

Relevant identified uses:

The SARS-CoV-2 Antigen Rapid Test is a rapid chromatographic immunoassay for the qualitative detection of SARS-CoV-2 nucleocapsid antigens in nasal and nasopharyngeal swab specimens. The SARS-CoV-2 Antigen Rapid Test is for professional in vitro diagnostics use only.

Uses advised against:

None.

1.3 Details of the supplier of the safety data sheet

Manufacturer:

Name: ACON Biotech (Hangzhou) Co., Ltd.

Address: No.210 Zhenzhong Road,
West Lake District, Hangzhou,
P.R. China, 310030

Phone: +86 571 87 96 35 69

E-mail: info@aconlabs.com

Authorized Representative in the EU:

Name: MedNet GmbH

Address: Borkstrasse 10
48163 Muenster, Germany

Phone: +49 251 32266-0

1.4 Emergency telephone number: +49 030/19240

SECTION 2: HAZARDS IDENTIFICATION

2.1 Classification of substance or mixture

This product does not meet the criteria for classification in any hazard class according to Regulation (EC) No 1272/2008 on classification, labelling and packaging of substances and mixtures.

2.2 Label elements

The product does not need to be labelled according to Regulation (EC) No. 1272/2008.

2.3 Other Hazards

The product does not contain any substance that meet the criteria for PBT/vPvB according to Annex XIII of Regulation (EC) No. 1907/2006.

SECTION 3: COMPOSITION /INFORMATION ON INGREDIENTS

3.1 Substance

Not Applicable.

3.2 Mixtures

3.2.1 Hazardous ingredients in Test Cassette

As per the Regulation (EC) No 1907/2006, the cassette and control swab is defined as an “Article” for which an SDS is not legally required. Thus, no substance need to be listed in this Section.

3.2.2 Hazardous ingredients in Buffer:

Extraction Buffer solution is accompanied with the SARS-CoV-2 Antigen Rapid Test in the kit box. Then concentration of the hazardous ingredients in the buffer is shown in below table:

Components	CAS number	Concentration	Classification according to Regulation (EC) No. 1278/2008 (CLP)	Specific Concentration. Limits, M-factors
Sodium azide	26628-22-8	0.02%	Acute Tox. 2 * (H300) Aquatic Acute 1 (H400) Aquatic Chronic 1 (H410)	N/A

SECTION 4: FIRST AID MEASURES

4.1 Description of first aid measures

If INHALATION: Move to fresh air. If not breathing, give artificial respiration. Do not use mouth-to-mouth method if victim ingested or inhaled; give artificial respiration with the aid of a pocket mask equipped with a one-way valve or other proper respiratory medical device. Immediate medical attention is required.

If SKIN Contact: Take off immediately all contaminated clothing. Wash off immediately with plenty of water for at least 15 minutes. Immediate medical attention is required.

If EYE Contact: Immediately flush eyes with plenty of water for at least 15 minutes. Assure adequate flushing by separating the eyelids with fingers. Get medical attention immediately.

If INGESTION: Clean mouth with water. Do not induce vomiting. Risk of aspiration! Keep airways free. Pulmonary failure possible after aspiration of vomit. Call a physician or Poison Control Center immediately.

4.2 Most important symptoms and effects, both acute and delayed

Symptoms/effects after skin contact: May cause skin irritation, corrosion and dermatitis. Drying-out effect resulting in rough and chapped skin.

Symptoms/effects after eye contact: May cause eye damage and corneal clouding.

Symptoms/effects after ingestion: May cause vomit.

4.3 Indication of any immediate medical attention and special treatment needed

No data available.

SECTION 5: FIREFIGHTING MEASURES

5.1 Extinguishing media

Use water spray, dry chemical or carbon dioxide.

5.2 Special hazards arising from the substance or mixture

No data available.

5.3 Advice for firefighters

Wear protective eyewear, gloves and clothing. Ensure self-safety.

SECTION 6: ACCIDENTAL RELEASE MEASURES

6.1 Personal precautions, protective equipment and emergency procedures

Not applicable.

6.2 Environmental precautions

Dispose the tests as medical rubbish.

6.3 Methods and material for containment and cleaning up

Dispose the tests as medical rubbish.

6.4 Reference to other sections

None.

SECTION 7: HANDLING AND STORAGE

7.1 Precautions for safe handling

Wear suitable laboratory coat and gloves. Avoid contacting with skin, eyes and mucous membranes. Take care not to splash, spill or splatter the buffer. Do not eat, drink or smoke in laboratory areas. Do not pipette the buffer by mouth. Wash hands and remove contaminated clothing after use.

7.2 Conditions for safe storage, including any incompatibilities

Store in the sealed package either at room temperature or refrigerated (2-30°C) and keep out of direct sunlight to ensure the product quality.

7.3 Specific end use(s)

No specific uses.

SECTION 8: EXPOSURE CONTROLS/PERSONAL PROTECTION

8.1 Control parameters

8.1.1 Occupational Exposure Limit Values:

Substance:	Sodium azide				
CAS No.	26628-22-8				
Country	Limit Value-Eight hours		Limit Value-Short term		Legal basis
	ppm	mg/m³	ppm	mg/m³	
Belgium		0.1		0.3	Data from GESTIS

Denmark		0.1		0.2	Database
European Union		0.1		0.3 (1)	
Finland		0.1		0.3 (1)	
France		0.1		0.3	
Germany (AGS)		0.2		0.4 (1)	
Germany (DFG)		0.2 inhalable aerosol		0.4 inhalable aerosol	
Hungary		0.1		0.3	
Ireland		0.1		0.3 (1)	
Italy		0.1		0.3	
Latvia		0.1		0.3 (1)	
Poland		0.1		0.3	
Spain		0.1		0.3	
Sweden				0.29 (1)	
Switzerland		0.2 inhalable aerosol		0.4 inhalable aerosol	
The Netherlands		0.1		0.3	
Turkey		0.1		0.3 (1)	
United Kingdom		0.1		0.3	
	Remarks				
European Union	Bold-type: Indicative Occupational Exposure Limit Values and Limit Values for Occupational Exposure Binding Occupational Exposure Limit Value - BOELV ~ (1) 15 minutes average value				
Finland	(1) 15 minutes average value				
France	Bold type: Restrictive statutory limit values				
Germany (AGS)	(1) 15 minutes average value				
Germany (DFG)	STV 15 minutes average value				
Ireland	(1) 15 minutes reference period				
Italy	skin				
Latvia	(1) 15 minutes average value				
Spain	Skin				
Sweden	(1) Ceiling Limit value				
Turkey	(1) 15 minutes average value				

8.1.2 Biological Limit Values:

No data available.

8.1.3 Monitoring Methods:

No data available.

8.2 Exposure controls

8.2.1 Appropriate engineering controls:

Use with adequate ventilation.

8.2.2 Personal protective equipment:

Use with adequate ventilation.

Eye/face protection: Not applicable.

Skin protection:

Hand protection: Not applicable.

Body protection: Not applicable.

Respiratory protection: Not applicable.

Thermal hazards: Not applicable.

8.2.3 Environmental exposure controls:

Do not allow to enter into surface water or drains.

SECTION 9: PHYSICAL AND CHEMICAL PROPERTIES

9.1 Information on basic physical and chemical properties

The below data applies to the buffer solution:

Appearance	colorless Liquid
Odor	odorless
Odor threshold	No data available
pH	8.0~9.0
Melting point/freezing point	No data available
Initial boiling point and boiling range	No data available
Flash point	No data available
Evaporation rate	No data available
Flammability (solid, gas)	No data available
Upper/lower flammability or explosive limits	No data available
Vapor pressure	No data available
Vapor density	No data available
Relative density	No data available
Solubility (ies)	No data available
Partition coefficient: n-octanol/water	No data available
Auto-ignition temperature	No data available
Decomposition temperature	No data available
Viscosity	No data available
Explosive properties	No data available
Oxidising properties	No data available

9.2 Other information

No data available.

SECTION 10: STABILITY AND REACTIVITY

10.1 Reactivity

Sodium azide (CAS No. 26628-22-8)	
Reaction	No data available.

10.2 Chemical stability

No known instability under normal conditions of use or storage.

10.3 Possibility of hazardous reactions

No data available.

10.4 Conditions to avoid

Keep away from open flames, hot surfaces and sources of ignition. Avoid dust formation.

10.5 Incompatible material

Acids, Oxidizing agents, Peroxides, Acid chlorides, Metals.

10.6 Hazardous decomposition products

Nitrogen oxides (NO_x), Sodium oxides, Carbon monoxide (CO), Carbon dioxide (CO₂).

SECTION 11: TOXICOLOGICAL INFORMATION

11.1 Information on toxicological effects

Acute toxicity

Sodium azide (CAS No. 26628-22-8)	
LD ₅₀ Oral (Mouse)	27 mg/kg
LC ₅₀ Inhalation (Rats)	0.054 and 0.52 mg/L
LD ₅₀ Dermal (Rabbits)	500-1000mg/kg

Skin corrosion/irritation	No data available.
Serious eye damage/irritation	No data available.
Respiratory or skin sensitization	No data available.
Germ cell mutagenicity	No data available.
Carcinogenicity	No component in this product is confirmed carcinogenicity by ACGIH, IARC, NTP or OSHA.
Reproductive toxicity	Sodium azide has a drastically toxic effect on the in vitro growth of mouse embryos at concentrations of 10 ⁻⁴ mol/L in the petri dish or greater.
STOT-single exposure	No data available.
STOT-repeated exposure	No data available.
Aspiration hazard	No data available.

SECTION 12: ECOLOGICAL INFORMATION

12.1 Toxicity

Sodium azide (CAS No. 26628-22-8)	
LC ₅₀ (Fish 1)	0.7 mg/L (96h, Lepomis macrochirus)
LC ₅₀ (Fish 2)	5.46 mg/L (96h, flow-through (Pimephales promelas)
LC ₅₀ (Fish 3)	0.8 mg/L (96h, Oncorhynchus mykiss)

12.2 Persistence and degradability

Sodium azide (CAS No. 26628-22-8)	
Persistence and degradability	Soluble in water Persistence is unlikely based on information available.

12.3 Bioaccumulative potential

Sodium azide (CAS No. 26628-22-8)	
Bioaccumulative potential	No data available.

12.4 Mobility in soil

Sodium azide (CAS No. 26628-22-8)	
Mobility in soil	Will likely be mobile in the environment due to its water solubility.

12.5 Results of PBT and vPvB assessment

This product does not contain any substances that are assessed to be PBT or vPvB.

12.6 Other adverse effects

No data available

SECTION 13: DISPOSAL CONSIDERATIONS

13.1 Waste treatment methods**Product**

Dispose as medical rubbish after being used

Contaminated packaging

Disposal should be in accordance with local, state or national legislation. Contaminated packaging must be disposed of in the same manner as the product.

SECTION 14: TRANSPORT INFORMATION

14.1 UN number

This product is not regulated for transport.

14.2 UN proper shipping name

This product is not regulated for transport.

14.3 Transport hazard class (es)

This product is not regulated for transport.

14.4 Packing group

This product is not regulated for transport.

14.5 Environmental hazards

No data available.

14.6 Special precautions for user

No data available.

14.7 Transport in bulk according to Annex II of MARPOL and the IBC Code

No data available.

SECTION 15: REGULATORY INFORMATION

15.1 Safety, health and environmental regulations/legislation specific for the substance or mixture

Not data available.

15.2 Chemical safety assessment

No data available.

SECTION 16: OTHER INFORMATION

16.1 Indication of Changes:

Version 1 Revision 0: First version, document in accordance with requirements for safety data sheets introduced by Regulation (EC) No 1907/2006 (REACH).

Version 2 Revision 0: Correct the PH value from “8.0” to “8.0~9.0” in section 9.1.

Version 3 Revision 0: Update the Relevant identified uses in section 1.2 to add the specimen type of “nasopharyngeal swab”.

16.2 Abbreviations and acronyms:

Acute Tox. 2: Acute Toxicity, Category 2

Aquatic Acute 1: Hazard to the aquatic environment – Acute, category 1

Aquatic Chronic 1: Hazard to the aquatic environment – Chronic, category 1

PBT: Persistent, Bioaccumulative and Toxic;

vPvB: Very Persistent and Very Bioaccumulative

16.3 Classification and procedure used to derive the classification for mixtures according to Regulation (EC) No 1272/2008 (CLP):

The product is not classified as a hazard mixture as per Regulation (EC) No 1272/2008 (CLP).

16.4 Relevant H-statements (number and full text):

H300 Fatal if swallowed.

H400 Very toxic to aquatic life.

H410 Very toxic to aquatic life with long lasting effects.

16.5 Further information

This information is based upon the present state of our knowledge.

This SDS has been compiled and is solely intended for this product.

Clinical Study Report
for
***Flowflex* SARS-CoV-2 Antigen**
Rapid Test

I. Intend for Use

The Flowflex SARS-CoV-2 Antigen Rapid Test is a lateral flow chromatographic immunoassay for the qualitative detection of the nucleocapsid protein antigen from SARS-CoV-2 in nasal swab specimens directly from individuals who are suspected of COVID-19 by their healthcare provider. The Flowflex SARS-CoV-2 Antigen Rapid Test does not differentiate between SARS-CoV and SARS-CoV-2.

II. Objective

A multi-site clinical study was conducted in China and USA to evaluate the performance of the Flowflex SARS-CoV-2 Antigen Rapid Test when compared to RT-PCR method.

III. Clinical Study Site and Study Period

Clinical Study Sites in USA:

Sample collection sites in USA	Testing sites in USA
<u>Site 1:</u> Boca Raton 6877 SW 18th Street Boca Raton, FL 33433	<u>Site 1:</u> Dr. Fowl 7200 Parkway drive, Suite 117, La Mesa, CA91942
<u>Site 2:</u> COVID CLINIC Westminster (WM) 2109 Westminster Mall Westminster, CA 92683	<u>Site 2:</u> COVID CLINIC Westminster (WM) 2109 Westminster Mall Westminster, CA 92683
<u>Site 3:</u> COVID CLINIC La Mesa (LM) 5601 Grossmont Center Drive La Mesa, CA 91942	<u>Site 3:</u> COVID CLINIC La Mesa (LM) 5601 Grossmont Center Drive La Mesa, CA 91942
<u>Site 4:</u> COVID CLINIC Down Town San Diego (DTSD) 1350 Third Avenue San Diego - San Diego County	<u>Site 4:</u> COVID CLINIC Down Town San Diego (DTSD) 1350 Third Avenue San Diego - San Diego County

Clinical Study Sites in China:

Sample collection sites in China	Testing sites in China
<u>Site 1:</u> <u>Shenzhen CDC</u> No. 8 Longyuan Road, Nanshan District, Shenzhen, P.R. China	<u>Site 1:</u> <u>Shenzhen CDC</u> No. 8 Longyuan Road, Nanshan District, Shenzhen, P.R. China
<u>Site 2:</u> Adicon No.208 Zhenzhong Road, West Lake District, Hangzhou, Zhejiang, P.R. China	<u>Site 2:</u> Adicon No.208 Zhenzhong Road, West Lake District, Hangzhou, Zhejiang, P.R. China

Study Period

Study Initiation Date: September, 2020

Study Completion Date: December, 2020

IV. Study acceptance criteria

Total Sensitivity: $\geq 85\%$

Total Specificity: $\geq 98\%$

V. Study Procedure:

The clinical performance of the Flowflex SARS-CoV-2 Antigen Rapid Test was evaluated at four (4) investigational sites in U.S and two (2) investigational sites in China using a total of 605 nasal swab specimens collected from the patients at multiple sites in U.S and China.

5.1 Clinical Study in USA

Material:

- SARS-CoV-2 Antigen Rapid Test, Lot# 202009001
- Comparison method:

TaqPath COVID-19 Combo Kit, FDA authorized RT-PCR test for emergency use, manufactured by Thermo Fisher Scientific, Inc.

CDC 2019-nCoV RT-PCR, ABI 7500DX, FDA authorized RT-PCR test for emergency use

- Nasal swab samples from infected patients and non-infected patients

Procedure:

A total of 153 nasal swab specimens were collected from the patients at multiple sites in U.S. The patients presenting the COVID-19 like symptoms within 14 days of symptom onset at the collection sites are enrolled.

The nasal swabs were randomized and blinded tested by operators following product package insert.

A companion nasopharyngeal (NP) swab was also collected from the same patient and confirmed as positive or negative and validated with Ct counts by the FDA EUA RT-PCR as a comparator method.

Test results:

Candidate method		RT-PCR method		
		Negative	Positive	Total
Flowflex Test Results	Negative	52	3*	55
	Positive	1	97	98
	Total	53	100	153

***3 samples with PCR CT value 32.9-33**

Relative Sensitivity: 97.0% (95% CI: 91.2%-99.4%)

Relative Specificity: 98.1% (95% CI: 89.1%-99.9%)

Accuracy: 97.4% (95% CI: 93.2%-99.2%)

5.2 Clinical Study in China

Material:

- SARS-CoV-2 Antigen Rapid Test, Lot# 202009001
- RT-PCR, Novel Coronavirus (2019-nCoV) Nucleia Acid Diagnostic Kit (PCR-Fluorescence Probing), FDA authorized RT-PCR test for emergency use, manufactured by Sansure BioTech Inc.
- Nasal swab samples from infected patients and non-infected patients

Procedure:

A total of 452 nasal swab specimens were collected from the patients at multiple sites in China. The patients presenting the COVID-19 like symptoms within 14 days of symptom onset at the collection sites are enrolled.

The nasal swabs were randomized and blinded tested by operators following product package insert.

A companion nasopharyngeal (NP) swab was also collected from the same patient and confirmed as positive or negative and validated with Ct counts by the FDA EUA RT-PCR as a comparator method.

Also the RT-PCR test results were confirmed by the clinical diagnostic result. RT-PCR positive specimens were all from diagnosis of COVID-19 patients

and RT-PCR negative specimens were all from non COVID-19 patients.

Test results:

Candidate method		RT-PCR method		
		Negative	Positive	Total
Flowflex Test Results	Negative	381	2*	383
	Positive	1	68	69
	Total	382	70	452

***2 samples with PCR Ct value 34-35**

Relative Sensitivity: 97.1% (95% CI: 89.6%-99.8%)

Relative Specificity: 99.7% (95% CI: 98.4%-99.9%)

Accuracy: 99.3% (95% CI: 98.0%-99.9%)

5.3 Summary of combined clinical studies at all sites:

Candidate method		RT-PCR method		
		Negative	Positive	Total
Flowflex Test Results	Negative	433	5	438
	Positive	2	165	167
	Total	435	170	605

Relative Sensitivity: 97.1% (95% CI: 93.1%-98.9%)

Relative Specificity: 99.5% (95% CI: 98.2%-99.9%)

Accuracy: 98.8% (95% CI: 97.6%-99.5%)

5.4 Positive results to be reported by different Ct value range

Ct value	RT-PCR Positive (+)	Proportion	Flowflex SARS-CoV-2 Antigen Rapid Test Positive (+)	PPA
≤27	86	50.6%	86	100%
27-30	38	22.4%	38	100%
>30-33	29	17.1%	27	93.1%
>33	9	5.3%	6	66.7%

Note: There are eight samples only have the PCR result of positive and no Ct value available.

Comparing with RT-PCR, the positive percent agreement (PPA) of the Flowflex SARS-CoV-2 Antigen Rapid Test is 100% for samples with Ct value ≤30, 93.1% for samples with Ct value from >30 to 33. For samples with Ct value >33, the PPA is 66.7%.

5.5 Positive results to be reported by days since symptom onset

Days Since Symptom Onset	RT-PCR Positive (+)	Proportion	Flowflex SARS-CoV-2 Antigen Rapid Test Positive (+)	PPA
0-3	81	46.3%	80	98.8%
4-7	62	37.0%	60	96.8%
>7	19	11.7%	17	89.5%

Note: There are four patients is asymptomatic individuals. And there are four patients lack "Days Since Symptom Onset" information.

Nasal swab specimens obtained early (≤ 7 days) after symptom onset may contain higher viral concentration.

5.6 Patient Demographics

Age Group	Total	RT-PCR Positive (+)	Flowflex SARS-CoV-2 Antigen Rapid Test Positive (+)	PPA
Children (Age < 18)	13	12	11	91.7%
Adult (Age 18 to 60)	565	132	128	97.0%
Elderly (Age ≥ 60)	23	22	22	100%

Note: There are four patients lack age information.

VI. Conclusions:

Using a total of 605 specimens tested at multiple sites in U.S and China, the Flowflex SARS-CoV-2 Antigen Rapid Test has sensitivity of 97.1%, specificity of 99.5%, and accuracy of 98.8% when comparing with FDA EUA RT-PCR.

*Clinical study data was collected from USA and China. Data analysis was performed by Azure Institute.



12/20/2020

Azure Institute
10125 Mesa Rim Road,
San Diego, CA 92130, USA



Clinical Study Report
for
Flowflex SARS-CoV-2 Antigen
Rapid Test

I. Intend for Use

The Flowflex SARS-CoV-2 Antigen Rapid Test is a lateral flow chromatographic immunoassay for the qualitative detection of the nucleocapsid protein antigen from SARS-CoV-2 in nasal swab and nasopharyngeal specimens directly from individuals who are suspected of COVID-19 by their healthcare provider. The Flowflex SARS-CoV-2 Antigen Rapid Test does not differentiate between SARS-CoV and SARS-CoV-2.

II. Objective

A multi-site clinical study was conducted in China and USA to evaluate the performance of the Flowflex SARS-CoV-2 Antigen Rapid Test when compared to RT-PCR method.

III. Clinical Study Site and Study Period

Clinical Study Sites in USA:

Sample collection sites in USA	Testing sites in USA
<u>Site 1:</u> Boca Raton 6877 SW 18th Street Boca Raton, FL 33433	<u>Site 1:</u> Dr. Fowl 7200 Parkway drive, Suite 117, La Mesa, CA91942
<u>Site 2:</u> COVID CLINIC Westminster (WM) 2109 Westminster Mall Westminster, CA 92683	<u>Site 2:</u> COVID CLINIC Westminster (WM) 2109 Westminster Mall Westminster, CA 92683
<u>Site 3:</u> COVID CLINIC La Mesa (LM) 5601 Grossmont Center Drive La Mesa, CA 91942	<u>Site 3:</u> COVID CLINIC La Mesa (LM) 5601 Grossmont Center Drive La Mesa, CA 91942
<u>Site 4:</u> COVID CLINIC Down Town San Diego (DTSD) 1350 Third Avenue San Diego - San Diego County	<u>Site 4:</u> COVID CLINIC Down Town San Diego (DTSD) 1350 Third Avenue San Diego - San Diego County
<u>Site 5:</u> Dr. Comunale's Office (DC) 502	<u>Site 5:</u> Dr. Comunale's Office (DC) 502

Euclid Ave #205 National City, CA 91950	Euclid Ave #205 National City, CA 91950
Site 6: Dr. Jonathan Leibowitz office 1343 55th street Brooklyn, NY 11219	Site 6: Dr. Jonathan Leibowitz office 1343 55th street Brooklyn, NY 11219
Site 7: Dr. Aline Benjamin office 624 Kings Highway Brooklyn, NY 11223	Site 7: Dr. Aline Benjamin office 624 Kings Highway Brooklyn, NY 11223

Clinical Study Sites in China:

Sample collection sites in China	Testing sites in China
<u>Site 1:</u> <u>Shenzhen CDC</u> No. 8 Longyuan Road, Nanshan District, Shenzhen, P.R. China	<u>Site 1:</u> <u>Shenzhen CDC</u> No. 8 Longyuan Road, Nanshan District, Shenzhen, P.R. China
<u>Site 2:</u> Adicon No.208 Zhenzhong Road, West Lake District, Hangzhou, Zhejiang, P.R. China	<u>Site 2:</u> Adicon No.208 Zhenzhong Road, West Lake District, Hangzhou, Zhejiang, P.R. China

Study Period

Study Initiation Date: September, 2020

Study Completion Date: May, 2021

IV. Study acceptance criteria

Total Sensitivity: $\geq 85\%$

Total Specificity: $\geq 98\%$

V. Study Procedure:

5.1 Procedure 1 for nasal swab specimens:

5.1.1 Material:

- SARS-CoV-2 Antigen Rapid Test, Lot# 202009001

- Comparison method:

TaqPath COVID-19 Combo Kit, FDA authorized RT-PCR test for emergency use, manufactured by Thermo Fisher Scientific, Inc.

CDC 2019-nCoV RT-PCR, ABI 7500DX, FDA authorized RT-PCR test for emergency use

RT-PCR, Novel Coronavirus (2019-nCoV) Nucleia Acid Diagnostic Kit (PCR-Fluorescence Probing), FDA authorized RT-PCR test for emergency use, manufactured by Sansure BioTech Inc.

- Nasal swab samples from infected patients and non-infected patients

5.1.2 Procedure:

A total of 882 nasal swab specimens were collected from the patients at multiple sites in USA and China. The patients presenting the COVID-19 like symptoms within 14 days of symptom onset at the collection sites are enrolled.

The nasal swabs were randomized and blinded tested by operators following product package insert.

A companion nasopharyngeal (NP) swab was also collected from the same patient and confirmed as positive or negative and validated with Ct counts by the FDA EUA RT-PCR as a comparator method.

5.1.3 Test results:

Candidate method		RT-PCR method		
		Negative	Positive	Total
Flowflex Test Results	Negative	587	8	595
	Positive	3	284	287
	Total	590	292	882

Relative Sensitivity: 97.3% (95% CI: 94.6%-98.7%)

Relative Specificity: 99.5% (95% CI: 98.4%-99.9%)

Accuracy: 98.8% (95% CI: 97.8%-99.3%)

5.1.4 Positive results to be reported by different Ct value range

Ct value	RT-PCR Positive (+)	Proportion	Flowflex SARS-CoV-2 Antigen Rapid Test Positive (+)	PPA
≤27	163	55.8%	163	100%
27-30	55	18.8%	55	100%
>30-33	47	16.1%	45	95.7%
>33	19	6.5%	13	68.4%

Note: There are eight samples only have the PCR result of positive and no Ct value available.

Comparing with RT-PCR, the positive percent agreement (PPA) of the Flowflex SARS-CoV-2 Antigen Rapid Test is 100% for samples with Ct value ≤ 30 , 95.7% for samples with Ct value from >30 to 33. For samples with Ct value >33 , the PPA is 68.4%.

5.1.5 Positive results to be reported by days since symptom onset

Days Since Symptom Onset	RT-PCR Positive (+)	Proportion	Flowflex SARS-CoV-2 Antigen Rapid Test Positive (+)	PPA
0-3	152	52.1%	150	98.7%
4-7	87	29.8%	85	97.7%
>7	23	7.9%	20	87.0%
Asymptomatic individuals	11	3.8%	10	90.9%

Note: There are 19 patients lack “Days Since Symptom Onset” information.

Nasal swab specimens obtained early (≤ 7 days) after symptom onset may contain higher viral concentration. The SARS-CoV-2 Antigen Rapid Test can also test specimens from asymptomatic individuals.

5.2 Procedure 2 for nasopharyngeal swab specimens:

5.2.1 Material:

- SARS-CoV-2 Antigen Rapid Test, Lot# 202009001
- Comparison method:

Hologic Panther SARS-CoV-2 (T000896), FDA authorized RT-PCR test for emergency use.

RT-PCR, Novel Coronavirus (2019-nCoV) Nucleia Acid Diagnostic Kit (PCR-Fluorescence Probing), FDA authorized RT-PCR test for emergency use, manufactured by Sansure BioTech Inc.

- Nasopharyngeal swab specimens from infected patients and non-infected patients

5.2.2 Procedure:

A total of 299 nasopharyngeal swab specimens were collected from the patients at multiple sites in UAS and China. The patients presenting the

COVID-19 like symptoms within 14 days of symptom onset at the collection sites are enrolled.

The nasopharyngeal swabs were randomized and blinded tested with SARS-CoV-2 Antigen Rapid Test by operators following product package insert. And all the specimens were also confirmed with an EUA RT-PCR as a comparator method.

5.2.3 Test results:

Candidate method		RT-PCR method		
		Negative	Positive	Total
Flowflex Test Results	Negative	175	3	178
	Positive	1	120	121
	Total	176	123	299

Relative Sensitivity: 97.6% (95% CI: 92.8% to 99.5%)

Relative Specificity: 99.4% (95% CI: 96.5% to 99.9%)

Accuracy: 98.7% (95% CI: 96.5% to 99.6%)

5.2.4 Positive results to be reported by different Ct value range

Ct value	RT-PCR Positive (+)	Proportion	Flowflex SARS-CoV-2 Antigen Rapid Test Positive (+)	PPA
≤27	58	47.2%	58	100%
>27-30	32	26.0%	32	100%
>30-33	21	17.1%	20	95.2%
>33	5	4.1%	3	60%

Note: There are 7 samples only have the PCR result of positive and no Ct value available.

Comparing with RT-PCR, the positive percent agreement (PPA) of the Flowflex SARS-CoV-2 Antigen Rapid Test is 100% for samples with Ct value ≤30, 95.2% for samples with Ct value from >30 to 33. For samples with Ct value >33, the PPA is 60%.

5.2.5 Positive results to be reported by days since symptom onset

Days Since Symptom Onset	RT-PCR Positive (+)	Proportion	Flowflex SARS-CoV-2 Antigen Rapid Test Positive (+)	PPA
0-3	41	33.3%	41	100%
4-7	60	48.8%	59	98.3%

>7	12	9.6%	10	83.3%
----	----	------	----	-------

Note: There are 10 patients lack “Days Since Symptom Onset” information. Specimens obtained early (≤ 7 days) after symptom onset may contain higher viral concentration.

5.3 Summary of combined clinical studies at all sites and with two sample types (nasal and nasopharyngeal swab specimens):

Candidate method		RT-PCR method		
		Negative	Positive	Total
Flowflex Test Results	Negative	762	11	773
	Positive	4	404	408
	Total	766	415	1181

Relative Sensitivity: 97.4% (95% CI: 95.3% to 98.6%)

Relative Specificity: 99.5% (95% CI: 98.6% to 99.9%)

Accuracy: 98.7% (95% CI: 97.9% to 99.3%)

VI. Conclusions:

Combine nasal swab specimens and nasopharyngeal swab specimens, a total of 1181 specimens tested at multiple sites in U.S and China, the Flowflex SARS-CoV-2 Antigen Rapid Test has sensitivity of 97.4%, specificity of 99.5%, and accuracy of 98.7% when comparing with FDA EUA RT-PCR.